



**Center for Clinical Standards and Quality/Quality, Safety & Oversight Group**

**Notice to Public of Limitation of CLIA Certificate**

**Date** December 19, 2023  
**Title** Notice to Public of Limitation of HENRY MAYO NEWHALL  
URGENT CARE, 05D2115911

The Centers for Medicare & Medicaid Services (CMS) is providing notice that on January 3, 2024, Henry Mayo Newhall Urgent Care's (23929 McBean PKWY, STE 102, Valencia, CA 91355-4467) CLIA certificate will be limited due to a subsequent occurrence of unsuccessful participation in proficiency testing for the analyte Cell I.D. or WBC Differential.

The facility has been notified that the CLIA certificate was limited on December 19, 2023. When a laboratory's CLIA certificate is limited, it is required to cease all patient iron laboratory testing, including waived and provider-performed microscopy testing regardless of whether or not the laboratory charges for the testing. CMS is committed to patient safety and quality of care.

The report of findings associated with this limitation action is available through CMS' Western and Central Branch Freedom of Information Officer. Please contact [CLIA\\_SanFrancisco\\_Office@cms.hhs.gov](mailto:CLIA_SanFrancisco_Office@cms.hhs.gov) for additional information.

As of August 18, 2017, public notice of revocations, suspension, and limitations will be posted on the Survey & Certification website at <https://www.cms.gov/Medicare/Provider-Enrollment-and-Certification/SurveyCertificationGenInfo/Termination-Notices.html>. For more information, please visit [www.medicare.gov](http://www.medicare.gov). To learn more about the termination process, please visit <https://www.cms.gov/Medicare/Provider-Enrollment-and-Certification/SurveyCertificationGenInfo/index.html>.