



CMS WILL NO LONGER BE PROVIDING PAPER COPIES OF HANDOUTS FOR THE MEETING. ELECTRONIC COPIES OF ALL MEETING MATERIALS WILL BE POSTED ON THE CMS WEBSITE PRIOR TO THE MEETING AT [HTTP://WWW.CMS.HHS.GOV/ICD9PROVIDERDIAGNOSTICCODES/03_MEETINGS.ASP](http://www.cms.hhs.gov/ICD9PROVIDERDIAGNOSTICCODES/03_MEETINGS.ASP)

DEPARTMENT OF HEALTH & HUMAN SERVICES
Centers for Medicare & Medicaid Services
7500 Security Boulevard
Baltimore, Maryland 21244-1850



Agenda

ICD-9-CM Coordination and Maintenance Committee
Department of Health and Human Services
Centers for Medicare & Medicaid Services
CMS Auditorium
7500 Security Boulevard
Baltimore, MD 21244-1850
ICD-9-CM Volume 3, Procedures
March 9 – March 10, 2011

Pat Brooks – Introductions and Committee overview
Co-Chairperson
March 9, 2011

9:00 AM – 5:30 PM ICD-9-CM Volume 3, Procedure presentations and public comments

Phone lines are available for participants who are unable to attend in person and who want to listen to the proceedings. Participants on the phone lines will be in “listen only” mode and will not be able to ask questions or provide comments. Phone participants should send any procedure code comments in writing to patricia.brooks2@cms.hhs.gov by April 1, 2011, the deadline for comments. We will **not** be posting an audio or written transcript of this meeting.

ICD-9-CM Topics

1. Cardiac Valve Replacement: Transcatheter Aortic.
Transapical Aortic, and Pulmonary
Pages 9-12
Ann B. Fagan
Craig R. Smith, MD
NY-Presbyterian Hospital
Doff B. McElhinney, MD
Dept. of Cardiology; Children's
Hospital, Boston, MA
2. PTCA/Atherectomy: Proposed revision of code 00.66
Pages 13-15
Ann B. Fagan
Jeffrey Chambers, MD
Cardiology Consultants, MN
3. Temporary Therapeutic Endovascular
Occlusion of Vessel
Pages 16-17
Ann B. Fagan
J. Neal Rutledge, MD
Austin Rad. Assoc.
4. Implanted Vascular Coils
Pages 18-19
Ann B. Fagan
Michael J. Alexander, MD
Dept. of Neurosurgery
Cedars-Sinai Med Ctr, LA, CA
5. Implantation of an Anti-Microbial Envelope
Pages 20-21
Pat Brooks
Daniel Lerner, MD
TyRx Pharma, Inc.
6. Implantable Ischemic Detection System
Pages 22-23
Ann B. Fagan
Kevin Bentley
Angel Medical
7. Insertion of Aqueous Drainage Shunt
Pages 24-25
Mady Hue
8. Four-Port Rechargeable Spinal Cord Neurostimulator
Pages 26-28
Mady Hue
Marshall Bedder M.D.
Dir. of Interventional Pain
Pacific Medical Centers
Seattle, Washington
9. Cardiac Lead Extraction
Pages 29-32
Pat Brooks
Laurence M. Epstein, MD
Brigham & Women's Hosp.

- | | |
|--|--|
| <p>10. Oxidized Zirconium Ceramic Hip Bearing Surface
Pages 33-34</p> | <p>Pat Brooks
Robert Bourne, MD
London Health Sciences Center</p> |
| <p>11. Sling Operation for Stress Incontinence
Pages 35-36</p> | <p>Celeste Beauregard</p> |
| <p>12. Sleeve Gastrectomy
Pages 37-38</p> | <p>Celeste Beauregard
Dan Duvall, MD, CMS</p> |
| <p>13. Electromagnetic Navigation Bronchoscopy
Pages 39-40</p> | <p>Celeste Beauregard
Alexander Chen, MD
Interventional Pulmonologist
Washington Univ. School
of Medicine and Barnes
Jewish Hospital</p> |
| <p>14. Ultrasound –enhanced Thrombolysis
Pages 41-43</p> | <p>Celeste Beauregard
Bradford A. Zakes
Cerevast Therapeutics</p> |
| <p>15. External Ventricular Drainage
Pages 44-46</p> | <p>Mady Hue
Dan Duvall, MD, CMS</p> |
| <p>16. Embolization of Uterine Artery
Pages 47-48</p> | <p>Mady Hue</p> |
| <p>17. Occlusion of Left Atrial Appendage
Pages 49-50</p> | <p>Ann B. Fagan
Adam Evan Saltman, MD, PhD
Interboro Surgical Associates, NY</p> |
| <p>18. Left Atrial Appendage Exclusion Femoral/Epicardial
Access
Pages 51-52</p> | <p>Ann B. Fagan
Randall Lee, MD, PhD
UCSF, CA</p> |
| <p>19. Ultrasonic Wound Debridement
Pages 53-54</p> | <p>Pat Brooks
Vickie R. Driver, MS, DPM,
FACFAS
Boston Univ. Medical Campus</p> |
| <p>20. Hydrosurgery/VersaJet
Pages 55-56</p> | <p>Pat Brooks
Christopher Attinger, MD</p> |



21. Non-excisional Debridement
Page 57

Georgetown Univ. Hospital

Mady Hue

22. Cerebral and Tissue Oximetry
Page 58-59

Celeste Beauregard
Harvey L. Edmonds, Jr., Ph.D.

23. Ultrasound Accelerated Thrombolysis
Pages 60-61

Celeste Beauregard
Douglas R. Hansmann, Ph.D

24. Addenda
Pages 62-64

Mady Hue

ICD-10 Topics:

1. Code Freeze
Page 65

Pat Brooks, CMS

2. Abbreviated Titles
Page 66

Pat Brooks, CMS

3. V28.0 ICD-10 MS-DRGs
Page 66

Pat Brooks, CMS

4. ICD-10-PCS Updates
Pages 67-68

Pat Brooks, CMS
Rhonda Butler, 3M

ICD-10-PCS Code Requests

1. Bearing Surface of Total Ankle Replacement
Page 69-71

Pat Brooks
Roger A. Mann, MD
Oakland Bone and Joint
Specialists

2. Interspinous Process Internal Fixation -
Dynamic stabilization vs. Static Distraction
Pages 72-73

Mady Hue
Reginald J. Davis, MD
Greater Balt. Neuro. Assoc.

3. Spinal Fusion
Page 74-75

Mady Hue

- | | |
|--|---|
| 4. Mesh (biological) in PCS
Pages 76-78 | Pat Brooks
Parag Bhanot, MD
Georgetown Univ. Hospital |
| 5. Intra-operative Monitoring PCS
Page79 | Mady Hue |
| 6. Cerebral and Tissue Oximetry
Pages 80-83 | Celeste Beauregard |

Registering for the meeting:

Information on registering online to attend the meeting can be found at:

<http://www.cms.hhs.gov/apps/events/>

For questions about the registration process, please contact Mady Hue at 410-786-4510 or marilu.hue@cms.hhs.gov.

Continuing Education Credits:

Continuing education credits may be awarded by the American Academy of Professional Coders (AAPC) or the American Health Information Management Association (AHIMA) for participation in CMS ICD-9-CM Coordination and Maintenance (C&M) Committee Meeting Conference Calls or on-site Meetings.

Continuing Education Information for American Academy of Professional Coders (AAPC)

If you have attended or are planning to attend a CMS ICD-9-CM Coordination and Maintenance (C&M) Committee Meeting Conference Call or on-site Meeting, you should be aware that CMS does not provide certificates of attendance for these. Instead, the AAPC will accept your e-mailed confirmation and call or meeting description as proof of participation. Please retain a copy of your e-mailed confirmation for these as the AAPC will request them for any conference call or meeting you entered into your CEU Tracker if you are chosen for CEU verification. Members are awarded one (1) CEU per hour of participation.

Continuing Education Information for American Health Information Management Association (AHIMA)

AHIMA credential-holders may claim 1 CEU per 60 minutes of attendance at an educational program. Maintain documentation about the program for verification purposes in the event of an audit. A program does not need to be pre-approved by AHIMA, nor does a CEU certificate need to be provided, in order to claim AHIMA CEU credit. For detailed information about AHIMA's CEU requirements, see the Recertification Guide on AHIMA's web site.



Please note: The statements above are standard language provided to CMS by the AAPC and the AHIMA. If you have any questions concerning either statement, please contact the respective organization, not CMS.

ICD-9-CM TIMELINE

A timeline of important dates in the ICD-9-CM process is described below:

March 9 – March 10 2011	ICD-9-CM Coordination and Maintenance Committee meeting.
April 1, 2011	There will not be any new ICD-9-CM codes implemented on April 1, 2011 to capture new technology.
April 1, 2011	Deadline for receipt of public comments on proposed code revisions discussed at the March 9-10, 2011 ICD-9-CM Coordination and Maintenance Committee meetings for implementation on October 1, 2011.
April 2011	Notice of Proposed Rulemaking to be published in the <u>Federal Register</u> as mandated by Public Law 99-509. This notice will include the final ICD-9-CM diagnosis and procedure codes for the upcoming fiscal year. It will also include proposed revisions to the DRG system on which the public may comment. The proposed rule can be accessed at: http://www.cms.hhs.gov/AcuteInpatientPPS/IPPS/list.asp
April 2011	Summary report of the Procedure part of the March 9, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be posted on CMS homepage as follows: http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes Summary report of the Diagnosis part of the March 10, 2011 ICD-9-CM Coordination and Maintenance Committee meeting report will be posted on NCHS homepage as follows: http://www.cdc.gov/nchs/icd9.htm
June 2011	Final addendum posted on web pages as follows: Diagnosis addendum at - http://www.cdc.gov/nchs/icd9.htm Procedure addendum at – http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes

July 15, 2011 Those members of the public requesting that topics be discussed at the September 14 – 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting must have their requests to CMS for procedures and NCHS for diagnoses.

August 1, 2011 Hospital Inpatient Prospective Payment System final rule to be published in the Federal Register as mandated by Public Law 99-509. This rule will also include all the final codes to be implemented on October 1, 2011.
This rule can be accessed at:
<http://www.cms.hhs.gov/AcuteInpatientPPS/IPPS/list.asp>

August 2011 Tentative agenda for the Procedure part of the September 14 – 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be posted on CMS homepage at -
<http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes>

Tentative agenda for the Diagnosis part of the September 14 – 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be posted on NCHS homepage at -
<http://www.cdc.gov/nchs/icd9.htm>

Federal Register notice for the September 14 –15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be published. This will include the tentative agenda.

August 12, 2011 On-line registration opens for the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting at:
<http://www.cms.hhs.gov/apps/events>

September 9, 2011 Because of increased security requirements, those wishing to attend the September 14 - 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting must register for the meeting online at:
<http://www.cms.hhs.gov/apps/events>

Attendees must register online by September 9, 2011; failure to do so may result in lack of access to the meeting.

September 14 –15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting.

Those who wish to attend the ICD-9-CM Coordination and Maintenance Committee meeting **must have registered for the meeting online by September 9, 2011**. You must bring an official form of picture identification (such as a drivers license) in order to be admitted to the building.

October 2011

Summary report of the Procedure part of the September 14 – 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be posted on CMS homepage as follows:
<http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes>

Summary report of the Diagnosis part of the September 14– 15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting report will be posted on NCHS homepage as follows:
<http://www.cdc.gov/nchs/icd9.htm>

October 1, 2011

New and revised ICD-9-CM codes go into effect along with DRG changes. Final addendum posted on web pages as follows:
Diagnosis addendum - <http://www.cdc.gov/nchs/icd9.htm>
Procedure addendum at -
<http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes>

October 07, 2011

Deadline for receipt of public comments on proposed code revisions discussed at the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee meetings for implementation on April 1, 2012.

November 2011

Any new ICD-9-CM codes required to capture new technology that will be implemented on the following April 1 will be announced. Information on any new codes to be implemented April 1, 2012 will be posted on the following websites:
<http://www.cms.hhs.gov/ICD9ProviderDiagnosticCodes>
<http://www.cdc.gov/nchs/icd9.htm>

November 18, 2011

Deadline for receipt of public comments on proposed code revisions discussed at the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee meetings for implementation on October 1, 2012.

Cardiac Valve Replacement: Transcatheter Aortic, Transapical Aortic, and Transcatheter Pulmonary

Issue: Recently developed transcatheter techniques allow certain heart valves to be replaced without conventional open heart surgery. Current ICD-9-CM codes do not clearly describe this method of valve replacement. Should new ICD-9-CM codes be established to distinctly identify these procedures?

New Technology Application?

Yes.

Food and Drug Administration (FDA) Approval? The SAPIEN transcatheter aortic valve began clinical trials in the United States in 2008. FDA submission was made in the fourth quarter of 2010 and approval is anticipated in late 2011. The CoreValve[®] transcatheter aortic valve is in the clinical trial process and other transcatheter aortic valve devices are also being developed.

The Melody[®] transcatheter pulmonary valve was approved by the FDA under a Humanitarian Device Exemption in January 2010.

Clinical Background: Conventionally, heart valves have been replaced via open heart surgery. Following a median sternotomy, a pericardotomy is performed and the heart is opened. The diseased native valve is surgically removed and the new valve is implanted in its place.

Valvotomy was once used as a less drastic alternative intervention to valve replacement but is rarely performed today. Due to the risks associated with open heart surgery in an elderly population with frequent major comorbid conditions, balloon valvuloplasty was introduced in the 1980's as a less invasive alternative to valve replacement. However, while it provides temporary relief, balloon valvuloplasty has not been shown to provide long-term benefit.

Recently, transcatheter heart valve replacement has emerged as a treatment option. The native valve is destroyed in situ and the new valve is implanted on top of its remains, replacing the native valve's structure and function.

The transcatheter valve assembly is mounted on a specially designed catheter for delivery to the implantation site. A multidisciplinary team of cardiologists and cardiac surgeons is usually required to perform a transcatheter valve replacement. The procedure typically requires between two to four hours to perform. Much of this OR time is associated with the high number of catheter exchanges and extreme precision needed to deliver the new valve and seat it properly, to ensure that it will take over full valve function immediately upon placement. Transcatheter valve replacement is performed in a specially equipped cardiac catheterization laboratory or

hybrid operating suite, with a backup cardiopulmonary bypass machine and full cardiac surgical team readily available should emergency surgical conversion be required.

Currently, transcatheter technique can be used to replace the aortic valve and the pulmonary valve.

There are two approaches to transcatheter aortic valve replacement: endovascular and transapical.

For aortic valve replacement using an endovascular approach, access is obtained through the femoral artery at the groin. A balloon valvuloplasty catheter is then advanced through the aorta and positioned over the diseased native aortic valve. Balloon valvuloplasty is performed. After removal of the valvuloplasty catheter, the delivery catheter is positioned across the native valve and the new bioprosthetic valve is expanded in place, crushing the native valve beneath it.

In the transapical approach, a small thoracotomy is made in the left 5th or 6th intercostal space. The apex of the heart is identified at the base of the left ventricle and opened. Guidewires and catheters are then advanced up through the left ventricle to reach the diseased native aortic valve. Balloon valvuloplasty is performed. After the valvuloplasty catheter is removed, the delivery catheter is positioned across the native valve and the new bioprosthetic valve is expanded in place, crushing the native valve beneath it. The opening at the apex of the heart is surgically repaired on the way out.

Transcatheter Pulmonary Valve Replacement: Transcatheter pulmonary valve replacement is typically performed in patients with certain congenital heart anomalies, such as pulmonary atresia, where a conduit can be constructed on the outside of the heart bypassing the diseased pulmonary valve. In these instances the valve being replaced is not at its normal anatomic location; the pulmonary valve is within the previously constructed right ventricle-to-pulmonary artery conduit.

For pulmonary valve replacement, the approach is endovascular. Access is typically obtained via the femoral vein at the groin. A catheter is then advanced into the right ventricle. From here, it is advanced into the previously created conduit and positioned over the existing valve within the conduit. Balloon valvuloplasty is performed, then the delivery catheter is positioned across the diseased existing valve and the new bioprosthetic valve is expanded in place, crushing the existing valve beneath it. When placed in pediatric patients, these valves do not grow with the patient and must be replaced from time to time.

CMS Recommendation

CMS recommends Option 2, as shown above.

Interim Coding

Assign transcatheter aortic valve replacement to code 35.21, Replacement of aortic valve with tissue graft.

Assign transapical aortic valve replacement to code 35.21, Replacement of aortic valve with tissue graft.

Assign transcatheter pulmonary valve replacement to code 35.25, Replacement of pulmonary valve with tissue graft.

PTCA/Atherectomy: Proposed Revision of Code 00.66

Issue: Coronary angioplasty and coronary atherectomy procedures are not synonymous and coronary angioplasty may be performed without an atherectomy. However, with both procedures included under one code, it is impossible to capture statistical information on utilization and outcomes. Should the codes be split out of existing code 00.66, Percutaneous transluminal coronary angioplasty [PTCA] or coronary atherectomy?

New Technology Application? No.

Food & Drug Administration (FDA) Approval: Not applicable.

Background: Angioplasty: Percutaneous transluminal coronary angioplasty (PTCA) is a minimally invasive catheter-based procedure to open up blocked coronary arteries, allowing blood to circulate unobstructed to the heart muscle. Local anesthesia is injected into the groin area and a needle is inserted into the femoral artery. A guide wire is placed through the needle and the needle is removed. An introducer is then placed over the guide wire, after which the wire is removed. A different sized guide wire is put in its place.

Next, a diagnostic catheter is advanced through the introducer over the guide wire, into the blood vessel. This catheter is then guided to the aorta and the guide wire is removed. Once the catheter is placed in the opening of one the coronary arteries, the physician injects dye and takes an x-ray.

If a treatable blockage is noted, the first catheter is exchanged for a guiding catheter. Once the guiding catheter is in place, a guide wire is advanced across the area of stenosis/occlusion; a balloon catheter is then advanced to the area of stenosis/occlusion. The balloon is inflated for a few seconds to compress the plaque against the artery wall. Then the balloon is deflated.

The physician may repeat this a few times, each time pumping up the balloon a little more to widen the passage for the blood to flow through. This treatment may be repeated at each area of stenosis/occlusion in the coronary arteries. Once the compression has been performed, contrast media is injected and an x-ray is taken to check for any change in the arteries. Following this, the catheter is removed and the procedure is completed.

Following this procedure, an atherectomy or stent insertion may also be performed.

Atherectomy: Atherectomy is a minimally invasive catheter-based procedure to remove plaque from arteries, and is useful in cases where the plaque is very hard due to calcification, plaque has built up in a coronary artery bypass graft, or to remove other difficult blockages, thus opening up blocked coronary arteries and allowing blood to circulate unobstructed to the heart muscle.

At the beginning of the procedure, medications to control blood pressure, dilate the coronary arteries, and prevent blood clots are administered. The patient is awake but sedated. A needle is

inserted into an artery in the groin, leg, or arm. A guide wire is placed through the needle and the needle is removed. An introducer is then placed over the guide wire, after which the wire is removed.

Next, a diagnostic catheter is advanced through the introducer into the blood vessel. This catheter is then advanced to the aorta root and maneuvered to the opening of one of the coronary arteries. Once the catheter is placed in the opening of one the coronary arteries, the physician injects dye and takes an x-ray.

If a treatable blockage is noted, a guide wire is advanced across the area of stenosis/occlusion and an atherectomy catheter is advanced into the diseased arterial segment. On the tip of the catheter is either a high-speed rotating device ("burr"), or a sharp blade. The burr grinds the plaque into minute particles, while the blade shaves the plaque away. The cutting head is positioned against the plaque and activated, and the plaque is ground up or suctioned out. Multiple passes with the atherectomy catheter may be necessary to physically remove the plaque from the artery. The atherectomy catheter may require periodic removal from the artery to empty the collection chamber of plaque (e.g. directional atherectomy catheter). It is then reintroduced into the artery and the process repeated until the desired amount of plaque is removed. Depending on the size of the vessel, length and extent of disease, more than one atherectomy device may be needed to open the entire diseased segment. Once the plaque has been removed, contrast media is injected and an x-ray is taken to check for any change in the arteries. Following this, the catheter is removed and the procedure is completed.

Following this procedure, a balloon angioplasty or stent insertion may also be performed.

The types of mechanical atherectomy are rotational, directional, and transluminal extraction. Rotational atherectomy uses a high speed rotating shaver to grind up plaque. Directional atherectomy was the first type approved, but is no longer commonly used; it scrapes plaque into an opening in one side of the catheter. Transluminal extraction coronary atherectomy uses a device that cuts plaque off vessel walls and vacuums it into a bottle. It is used to clear bypass grafts. Excimer laser catheters may also be used to pulverize the plaque in a similar process referred to as laser atherectomy or laser angioplasty.

Atherectomy is performed in a cardiac catheterization lab. It can be used instead of, or along with, balloon angioplasty. Atherectomy is successful about 95% of the time. Plaque forms again in 20-30% of patients.

Current Coding:

Both percutaneous transluminal coronary angioplasty [PTCA] and percutaneous transluminal coronary atherectomy are coded to 00.66.

Coding Options:

Option 1: Make no change to this code. Continue coding both angioplasty and atherectomy using ICD-9-CM procedure code 00.66, Percutaneous transluminal coronary angioplasty [PTCA] or coronary atherectomy.

Option 2: Create a new code to describe atherectomy, as shown below:

	17.5	Additional Cardiovascular Procedures
New code	17.53	Transluminal coronary atherectomy
		Directional atherectomy
		Excimer laser atherectomy
		Rotational atherectomy
		That by laser
		That by percutaneous approach
		That by transluminal extraction
		Code also any transluminal coronary angioplasty (00.66)
	00.6	Procedures on Blood Vessels
Revise title	00.66	Percutaneous transluminal coronary angioplasty [PTCA] or
		<u>coronary atherectomy</u>
Add code also note		Code also any: transluminal coronary atherectomy (17.53)

CMS Recommendation:

Create a new code at 17.53, Transluminal coronary atherectomy, as described above.

Interim Coding:

Continue to use ICD-9-CM procedure code 00.66 to describe both percutaneous transluminal coronary angioplasty [PTCA] and percutaneous transluminal coronary atherectomy.



Temporary Therapeutic Endovascular Occlusion of Vessel

Issue: There are no existing ICD-9-CM procedure codes that adequately classify procedures in which the abdominal aorta is partially occluded via an endovascular balloon catheter. This treatment is for patients with cerebral ischemia.

New Technology Application? No.

Food & Drug Administration (FDA) Approval: The FDA approved the CoAxia NeuroFlo™ Catheter device under the Humanitarian Device Exemption (HDE) program. According to the FDA’s approval letter, “This device is indicated for the treatment of cerebral ischemia resulting from symptomatic vasospasm following aneurismal subarachnoid hemorrhage, secured by either surgical or endovascular intervention for patients who have failed maximal medical management.”

Background: According to the FDA web site, the NeuroFlo™ Catheter is a potential treatment for victims of ischemic stroke who have not responded to other forms of treatment. The catheter is inserted through the femoral artery and into the descending aorta where it uses balloons to partially restrict blood flow, diverting flow from the lower extremities to the cerebral collaterals. The assumption is that this may improve neurologic outcomes in these patients. Initial efficacy studies are still pending.

Current Coding: This procedure is a minimally invasive form of temporary vascular occlusion, the purpose of which is to prevent blood flow to an area of the body. There is no unique or predecessor code that currently describes this procedure. Code 39.79, Other endovascular procedures on other vessels, is currently the most appropriate place to identify this procedure.

Coding Options:

Option 1: Do not create a new code. Use code 39.79, Other endovascular procedures on other vessels, to identify temporary therapeutic endovascular occlusion of vessel. Add this language to the procedure Index and to the inclusion terms in the Tabular:

39.7	Endovascular Procedures on Vessel(s)
39.79	Other endovascular procedures on other vessels
Add inclusion term	<u>Temporary balloon catheter occlusion</u>
Add inclusion term	<u>Temporary therapeutic occlusion</u>

Option 2: Create a new code which specifically identifies this procedure.

39.7	Endovascular Procedures on Vessel(s)
New code	39.77 Temporary therapeutic endovascular occlusion of vessel That by balloon catheter

CMS Recommendation:

CMS recommends creation of a new code as specified in Option 2, above. If this option is selected, appropriate Tabular changes to existing code 39.79 would be made to exclude new code 39.77.

Interim Coding:

Use interim code 39.79, Other endovascular procedures on other vessels.

Insertion of Multiple Coils for the Embolization or Occlusion of Head or Neck Vessels

Issue: There are currently no ICD-9-CM procedure codes that identify the number of coils inserted during an embolization or occlusion procedure in vessels of the head or neck. Should codes be created to capture this level of detail?

New Technology Application? No

Food & Drug Administration (FDA) Approval: Not applicable.

Background: Aneurysms result when weakness in the wall of an artery causes ballooning or out-pouching of blood vessels. Untreated, intracranial aneurysms may rupture leading to hemorrhagic stroke. Ruptured brain aneurysms are devastating events that have an extremely poor prognosis, with a one-year mortality rate of 50% and an additional 30% of patients suffering permanent neurological and cognitive deficits. A recent study estimated the incidence of unruptured intracranial aneurysms at approximately 23,000 per year in the United States. Neurovascular coils for occluding intracranial aneurysms became available in the 1990s and are used for treating both non-ruptured and ruptured aneurysms. For the treatment to be effective, aneurysms are embolized with coils until dense packing is achieved. The distinguishing element of the procedure is the number of coils that are inserted by the physician to achieve this dense packing.

Once the location of the cerebral aneurysms is identified using angiography, the first coil is advanced through the microcatheter into the aneurysm. This coil may be repeatedly withdrawn and advanced to ensure proper positioning. The placement of the first coil is commonly referred to as “framing the aneurysm.” A subsequent coil is selected, then prepared and advanced into the microcatheter and then into the aneurysm. The coil may also be withdrawn and advanced to ensure proper positioning. This process is continued with each subsequent coil until the embolization is nearly complete. The second coil and each subsequent coil are often referred to as “the filling coils.” Finally, to complete the embolization of the aneurysm, one or more softer and shorter “finishing coils” are inserted.

Because the size and type of aneurysms varies considerably, the number of coils placed varies considerably too. Depending on the size of the aneurysm, the number of coils placed may range from 3 coils for simple aneurysms up to 70 coils for giant aneurysms. The placement of additional coils adds greatly to the clinical complexity of the procedure, particularly in terms of duration of the procedure, difficulty, and risk to the patient. The number of coils placed varies with the aneurysm's length, width, and shape, e.g. berry, elongated or bilobed. In stratifying the ranges of coils that can be used for embolization, aneurysms of lesser complexity may use 1 to 10 coils. Aneurysms of moderate complexity may use 11 to 25 coils. Aneurysms that use more than 25 coils are of the highest complexity.

It is not currently possible to relate the number of coils placed to clinical outcomes, long-term durability and effectiveness, and quality measures in the treatment of aneurysms because the level of data collected, today, lacks this element of specificity. Information on the number of coils is routinely available in the medical record. In addition to the procedure report, the implant record and implant stickers provide a precise count of coils implanted.

Current Coding: There is currently no way to report the number of coils used to occlude an aneurysm, therefore it is not coded.

Coding Options:

Option 1: Do not create a code or series of codes to capture this information.

Option 2: Create a series of adjunct codes to identify the number of coils inserted into an aneurysm.

- 00.9 Other procedures and interventions
- New code 00.9x Insertion of one to 10 vascular coils
Code also any fluoroscopy or angiogram
- New code 00.9x Insertion of 11 to 24 vascular coils
Code also any fluoroscopy or angiogram
- New code 00.9x Insertion of 25 or more vascular coils
Code also any fluoroscopy or angiogram

- 39.75 Endovascular embolization or occlusion of vessel(s) of head or neck using bare coils
Add code also note Code also the number of coils inserted (00.9x-00.9x)

- 39.76 Endovascular embolization or occlusion of vessel(s) of head or neck using bioactive coils
Add code also note Code also the number of coils inserted (00.9x-00.9x)

CMS Recommendation: CMS recommends the adoption of Option 1; do not create a code or codes to describe the number of coils inserted. There are many ways to capture the parameters of the procedure (number of coils inserted, exact size or shape of the aneurysm, exact location of the aneurysm, delivery system of the coils), and all are good from a study perspective. However we question the value added to the data base. We encourage all comments regarding this recommendation.

Interim Coding: Do not code.



Implantation of Antimicrobial Envelope

Issue: There is not a code that identifies whether or not an antimicrobial envelope is used with the insertion of a pacemaker or defibrillator. The antimicrobial envelopes are used in an effort to reduce post operative infections. The requestor states that a new ICD-9-CM procedure code could be used to monitor the occurrence of post operative infections following the use of an antimicrobial envelope.

New Technology Application? No.

Food and Drug Administration (FDA) Approval: AIGISRTM Antibacterial Envelope, manufactured by TYRX (Monmouth, New Jersey), received 510K clearance on January 16, 2008. This device is only intended to be used in conjunction with pacemakers and implantable defibrillators.

Background: With the widespread use of cardiovascular implantable electronic devices (CIED), such as pacemakers and implantable defibrillators, there has been an associated rise in complications, including infection. Fabric pouches, into which the generators are placed prior to insertion, have been used for several decades to reduce device migration while local and systemic prophylactic antibiotics have been used to reduce post implantation infection. A manufacturer has recently released a fabric pouch composed of fibers that contain embedded antibiotics. The antibiotics are released over a number of days following implantation in order to provide continuous local antibiotic activity in the immediate post surgical period. The manufacturer states that an ICD-9-CM procedure code would permit assessment of clinical outcomes in patients with and without an antimicrobial envelope.

Current Coding: Hospitals currently code the primary surgical procedure including the insertion of a cardiac pacemaker, implantable cardioverter-defibrillator, or cardiac resynchronization therapy device. No code is assigned to capture the use of supplies or adjuncts used during surgery, such as a traditional fabric pouch that does not elute antibiotics. If a new code were created for the use of the antibacterial envelope, coders would need to review the operative report to look for the use of the device.

Coding Options:

Option 1: Do not create a new code for the use of this envelope during surgery. Continue to code the insertion of the pacemaker or the defibrillator.

Option 2: Create a new code in subcategory 17.5, Additional cardiovascular procedures, as follows.

New code 17.5x Implantation of anti-microbial envelope

CMS Recommendation: Option 1. Do not create a new code. The use of supplies such as this envelope during part of a procedure is not commonly captured in ICD-9-CM codes.



Implantable Ischemic Detection System (IIDS)

Issue: The ICD-9-CM procedure codes do not have a unique code to identify an implantable ischemic detection system for the detection of coronary plaque rupture and ischemia. Should a distinct code be created?

New Technology Application?

No.

Food & Drug Administration (FDA) Approval:

The device is currently undergoing an IDE trial, with data anticipated for the second quarter of 2012. FDA clearance would likely be sought if at least one of the three efficacy endpoints is met.

Background: The AngelMed-Guardian® implantable ischemic detection system (IIDS) is designed to provide early detection and patient monitoring for ischemic events in ambulatory patients. The purpose of the IIDS is to provide a means of detecting rapidly progressive and significant ST shifts in EKGs and other cardiac irregularities, capture the related EEG data, and alert the patient to seek medical attention when a threshold has been met or exceeded.

Components of the device: a programmable implantable monitoring device (IMD) that is implanted in the same manner as a pacemaker, including a right ventricular lead and a lead adapter, a pager-sized portable external alarm device, and a workstation used to program the detection parameters and upload EEG information from the implant.

Procedure: the IMD is implanted under the skin in the left pectoral region in the same manner as a single chamber pacemaker. It attaches to a bipolar pacemaker lead inserted transvenously and is implanted into the apex of the right ventricle of the heart.

Current Coding: The implantable ischemic detection system is very similar to pacemaker implantation; therefore many existing codes can be utilized for this implantation. See the interim coding section of this background paper for specific codes already in existence.

Coding Options:

Option 1: Do not create a unique code for description of the implantable ischemic detection system. Implantation of this device is quite similar to implantation of a pacemaker, and ICD-9-CM procedure coding already exists for that procedure.

Option 2: Create new codes describing the implantable ischemic detection system for the codes that do not correspond to existing pacemaker codes, as follows:

17.5 Additional cardiovascular procedures

- New code 17.54 Insertion or replacement of implantable cardiovascular monitoring system
Implantable ischemic detection system (IIDS)
Total system
- New code 17.55 Insertion of cardiovascular implantable monitoring device (IMD)
Note: IMD refers to transceiver/battery component
Code also any insertion of lead(s) e.g. into ventricle (37.70, 37.71, 37.76)
- New code 17.56 Revision or removal of implantable monitoring device (IMD)
Note: IMD refers to transceiver/battery component
Removal without replacement of implantable monitoring device
(IMD)
Repair of implantable monitoring device (IMD)
Code also any removal of leads without replacement (37.77)

CMS Recommendation:

Option 1; do not create a new code or codes for this device. The device is currently in clinical trials, and does not have FDA approval at this time.

Interim Coding:

The implantable ischemic detection system is very similar to pacemaker implantation; therefore these existing codes can be used to describe portions of the implantation of this device.

- 37.70, Initial insertion of lead [electrode], not otherwise specified
- 37.71, Initial insertion of transvenous lead [electrode] into ventricle
- 37.75, Revision of lead [electrode]
- 37.76, Replacement of transvenous atrial and/or ventricular lead(s) [electrode]
- 37.77, Removal of lead(s) [electrode] without replacement
- 37.79, Revision or relocation of cardiac device pocket
- 37.81, Initial insertion of single-chamber device, not specified as rate responsive.

CMS Recommendation: We recommend option 2 as shown above.

Interim coding advice: Continue to assign code 12.69, Other fistulizing procedure, to identify the insertion of an aqueous drainage shunt.

Four-Port Spinal Cord Neurostimulator

Issue: Currently there is not a unique ICD-9-CM procedure code to identify a four-port spinal cord neurostimulator. Should a new code be created?

New Technology Application? No.

FDA Approval: The four-port neurostimulator system will be commercially available, subject to FDA approval, in the first or second quarter of 2012.

Background: Spinal cord stimulator (SCS) technology is used as a late or last resort therapy to treat patients with intractable chronic pain who are unresponsive to other more conservative treatments (e.g., pharmacologic therapy). SCS operates on a platform that includes electrode array(s) that contain four to eight electrodes (or contact points) and an implantable pulse generator (IPG). There are also external accessories that include a programmer to control the stimulation settings and, in the instances where the IPG is rechargeable, an external charging system to recharge the internal battery via radiofrequency signal. In 2004, this technology was revolutionized further with the introduction of rechargeable neurostimulator generators to better extend battery life in mitigating the amount of replacement surgeries compared to non-rechargeable generators.

The theory behind SCS therapy is that pulsed electrical stimulation to the spinal nerves innervating the area of pain will result in reduction in pain. The electrode arrays are implanted so that the contacts on the array are near the spinal cord and the arrays are attached to the IPG.

The IPG contains one or more “ports” or “channels” through which the electrode arrays are attached to the IPG. The earliest IPGs had a single port/channel while newer IPGs have two ports/channels. Traditionally each port is connected to a single electrode array. However, clinicians have found that it is sometimes necessary to implant more than two arrays in order to provide effective SCS therapy. This occurs in patients with multiple areas of pain or difficult to reach target areas.

Even with a two port IPG, it is impossible to have more than 16 active contact points. So, if it is medically necessary to implant three or more arrays to achieve optimal pain control (i.e., to have more than 16 contact points), clinicians have had to implant a second IPG. There are a number of technical and clinical drawbacks to implantation of two IPGs in a single patient. Therefore, in the last few years, clinicians have started to use “splitters” which are “Y”-extensions that allow two arrays to be attached to a single port/channel of a dual port IPG. This means that up to four arrays can be attached to a single port IPG. The limitation of splitters used in conjunction with an IPG is that only a maximum of 16 contact points can be active/programmed no matter how many arrays are attached to a dual port IPG. Unfortunately, the choice as to which contacts are active must be made at the time of implantation. This choice is permanent because it is not possible to make an active contact inactive or vice versa.

Technology

The new SCS system will include a rechargeable IPG with four ports/channels. This system will allow placement of up to four independent electrode arrays to a single IPG without the use of splitters. The availability of a four port/channel IPG is likely to replace the use of splitters because the connections of the arrays to the IPG will be more robust and technically easier.

Importantly, the new rechargeable four port/channel IPG will likely be used in a different patient population than the two port/channel IPGs. More specifically, a number of patients refractory to other treatments have pain that migrates from one site to another (e.g., from the buttock area to the lower leg or foot). In these patients, dual array IPGs are not optimal because of the variety of places pain can migrate and a second IPG may need to be implanted. IPGs with four ports are able to provide relief over a much wider area and an array can be easily added when necessary. Following are a few clinical scenarios:

- 1) Patients with severe disabling low back pain with multiple locations along the vertebral column to target for treatment;
- 2) Patients with bilateral back or lower extremity pain that currently requires two IPGs;
- 3) Patients with reflex sympathetic dystrophy syndrome (RSDS). Data indicate that SCS may be more effective than conventional treatments (Kemler, et. Al. 2002);
- 4) Patients with new areas of pain that develop after the initial neurostimulator generator implant without having to insert a second pulse generator.

Current Coding

The ICD-9-CM procedure codes currently used to report SCS therapy are:

86.94, Insertion or replacement of single array neurostimulator pulse generator, not specified as rechargeable

86.95, Insertion or replacement of dual array neurostimulator pulse generator, not specified as rechargeable

86.96, Insertion or replacement of other neurostimulator pulse generator

86.97, Insertion or replacement of single array rechargeable neurostimulator pulse generator

86.98, Insertion or replacement of dual array rechargeable neurostimulator pulse generator

Codes 86.95 and 86.98 can be (and are) used to report the insertion or replacement of a two port/channel IPG when splitters are used to allow implantation of four arrays. However, there is no code that specifically identifies an IPG with four ports/channels. According to the

requestor, the existing ICD-9-CM procedure codes do not appear to accurately identify the insertion or replacement of a four port/channel rechargeable neurostimulator generator.

Coding Options

Option 1: Do not create a new code. Use existing code 86.96, Insertion or replacement of other neurostimulator pulse generator, and add appropriate inclusion/exclusion term(s).

Option 2: Revise existing code 86.98, Insertion or replacement of dual array rechargeable neurostimulator pulse generator.

Revise code title 86.98 Insertion or replacement of ~~dual~~ multiple array (two or more) rechargeable neurostimulator pulse generator

Add inclusion term Rechargeable pulse generator (multiple array, multiple channel) for spinal cord neurostimulator

Option 3: Create a new code to identify a multiple array (more than two) rechargeable neurostimulator pulse generator. Add exclusion terms as appropriate.

New code 86.08 Insertion or replacement of multiple array (more than two) rechargeable neurostimulator pulse generator.

CMS Recommendation: CMS recommends option 2 as stated above.

Interim coding advice: Continue to use existing code 86.98, Insertion or replacement of dual array rechargeable neurostimulator pulse generator.



Cardiac Lead Extraction

Issue: The ICD-9-CM procedure coding system does not differentiate between complex cardiac lead extractions versus simple lead extractions or lead abandonment (capping of leads). The requestor stated that the more complex lead extraction procedures require the assistance of specialized devices and utilization of additional time and resources to manage various clinical and mechanical complications that arise with implanted pacemaker and defibrillator (ICD) leads. A new procedure code was requested which would distinguish more resource-intensive complex lead extraction procedures from simple lead revision or removal procedures that require no special tools or resources. The requestor asserts that this added coding detail is necessary in order to better classify complex lead extraction procedures in registries and databases for tracking and monitoring outcomes and performance, identification of resource utilization, and for developing a uniform definition in the literature.

New Technology Application? No.

Food & Drug Administration (FDA) Approval: Yes

Background: The field of cardiac rhythm management has advanced considerably over the past two decades. Improvements in technology coincide with the rapidly increasing number of patients that require the implantation of pacemaker and defibrillator systems. The integrity of the implanted system is essential for appropriate device therapy. Unfortunately, a proportion of recipients experience some type of device system-related complications requiring a complex lead extraction and a replacement of a lead. In other simpler circumstances, physicians may choose to remove existing leads without any special tools or resources. The leads can also be capped and abandoned in the implant vein before adding a replacement lead. Resources required for complex lead extraction versus lead abandonment are quite different.

Lead Extraction and Clinical Indications: Clinical indications for lead extraction include infection, chronic pain, thrombosis or venous stenosis, and non-functional leads. As technology continues to advance, it is critical to have the ability to identify complications and accurately trace their origin. Implantable cardiac defibrillator (ICD) leads may require specialized extraction equipment. Extraction tools can be of different types, and include locking stylets and sheaths which vary in design and action.

The Procedure: Lead extraction can involve varying degrees of difficulty. The extraction can be performed in a variety of settings including an operative suite. Lead extraction may involve the use of one or more specialized tools to remove leads from the patient's major vein and/or heart. An implanted lead resides in the venous system and heart where over time, scar tissue encases the lead at several points and binds it to the anatomy, thereby preventing the lead from

simply sliding out upon attempted removal. Typical sites for scar tissue include the subclavian vein, innominate vein, superior vena cava, right atrium, and right ventricular apex.

Commonly, a locking stylet is inserted through the inner lumen of a lead to enable a physician to apply traction to the lead for removal. Often a lead extraction sheath is needed in addition to a locking stylet to manipulate the scar tissue to free the lead.

- Basic sheaths and more advanced rotating mechanical sheaths involve a mechanical cutting action to dissect the binding scar tissue.
- Powered sheaths involving laser energy or radiofrequency energy to assist the lead extraction process by vaporizing or ablating the scar tissue.
- Other specialized snaring tools and sheaths may be used to remove leads from locations other than the implant vein, such as the femoral vein.

In contrast, lead abandonment in this scenario involves no additional preparation beyond a standard defibrillator implant procedure. The malfunctioning lead is simply disconnected from the generator, and the terminal pin is covered with a plastic cap before being sutured down within the device pocket after the replacement lead is inserted.

Tracking Complications

Recording all complications is crucial for quality assessment and quality improvement. Documentation of complications is made difficult by the fact that several procedures may be performed on the patient in succession during the same or closely spaced hospitalizations. Examples of potential complications related to extraction or removal include:

Major Complications

- Death
- Cardiac avulsion or tear requiring thoracotomy, pericardiocentesis, chest tube, or surgical repair
- Vascular avulsion or tear (requiring thoracotomy, pericardiocentesis, chest tube, or surgical repair)
- Pulmonary embolism requiring surgical intervention
- Respiratory arrest or anesthesia related complication leading to prolongation of hospitalization
- Stroke
- Pacing system related infection of a previously non-infected site

Minor Complications

- Pericardial effusion not requiring pericardiocentesis or surgical intervention
- Hemothorax not requiring a chest tube
- Hematoma at the surgical site requiring reoperation for drainage

- Arm swelling or thrombosis of implant veins resulting in medical intervention
- Vascular repair near the implant site or venous entry site
- Hemodynamically significant air embolism
- Migrated lead fragment without sequelae
- Blood transfusion related to blood loss during surgery
- Pneumothorax requiring a chest tube
- Pulmonary embolism not requiring surgical intervention

Current Coding:

The following are the existing codes and descriptors for cardiac lead removal:

- 00.52 Implantation or replacement of transvenous lead [electrode] into left ventricular coronary venous system
- 37.76 Replacement of transvenous atrial and/or ventricular lead(s) [electrode]
- 37.77 Removal of lead(s) [electrode] without replacement
- 37.89 Revision or removal of pacemaker device
- 37.97 Replacement of automatic cardioverter/defibrillator lead(s) only

These codes are not differentiated as to the complexity of the patient management, procedural approach, and need for special tools or techniques to assist with the lead removals.

Complications leading to the need for removal of the leads are captured through a separately reported diagnosis code. Coders do not review the medical record to determine what types of tools are needed or used to perform a specific procedure such as a lead removal. Introducing such a new concept may pose problems for coders who rely on physician documentation of the procedure performed.

Coding Options:

Option 1: Do not create new codes for complex lead extractions. ICD-9-CM does not differentiate procedures based on tools that may be used to assist with the procedures. Continue to use the existing codes.

Option 2:

- | | |
|--------------------|---|
| Create new code | 37.XX Complex transvenous cardiac lead extraction, atrial and/or ventricular lead(s) [electrode] with device assistance (laser) (mechanical) (radiofrequency) |
| | 37.76 Replacement of transvenous atrial and/or ventricular leads(s) [electrode] |
| Add code also note | <u>Code also any: Complex transvenous cardiac lead extraction (37.XX)</u> |



00.52 Implantation or replacement of transvenous lead [electrode] into left ventricular coronary venous system
Add code also note Code also any: Complex transvenous cardiac lead extraction (37.XX)

37.97 Replacement of automatic cardioverter/defibrillator lead(s) only
Add code also note Code also any: Complex transvenous cardiac lead extraction (37.XX)

With this option new code 37.XX is created. The main code descriptors for the existing codes 37.76, 00.52, and 37.97 are not changed. The addition of subterms for the existing codes would provide instructional guidance for capturing complex lead extraction procedures by the use on new code 37.XX.

CMS Recommendation: Option 1. Do not create a new code for complex lead extractions.

Interim Coding: Continue to use existing codes as follows

- 00.52 Implantation or replacement of transvenous lead [electrode] into left ventricular coronary venous system
- 37.76 Replacement of transvenous atrial and/or ventricular lead(s) [electrode]
- 37.77 Removal of lead(s) [electrode] without replacement
- 37.89 Revision or removal of pacemaker device
- 37.97 Replacement of automatic cardioverter/defibrillator lead(s) only

Oxidized Zirconium Ceramic Hip Bearing Surface

Issue: There is not a unique code to capture the use of oxidized zirconium as part of the ceramic bearing surfaces in total hip arthroplasties. Oxidized zirconium is technically a layer of zirconia ceramic on a metal head and for hips is captured by code 00.77, Hip bearing surface, ceramic on polyethylene.

New Technology application? No.

Food & Drug Administration (FDA) Approval: The oxidized Zirconium articulating surface has been FDA approved since 1992 for hip joints and 1996 for knee joints.

Background: Advances in technology have produced a wide variety of bearing surfaces that may coexist in a single category in ICD-9-CM. Most bearing surfaces are based upon solid metal or solid ceramic surfaces abutting metal, ceramic or polyethylene surfaces. The choice of surface composition is based on the relative characteristics of the surface: The oxidizing process can cause certain metals to form a ceramic surface overlying the metal core, creating a metal bearing with a unique ceramic surface. One example is an articulating surface which is an oxidized biocompatible substrate with the toughness of metal and superior wear characteristics of ceramic without the brittleness.

- Metal's chief advantage is toughness and fracture resistance
- Metal's chief disadvantage may be wear characteristics and ion shedding
- Ceramic's chief advantage is superior wear characteristics
- Ceramic's chief disadvantage is brittleness
- Polyethylene is used to take advantages of articulation of a hard surface against a "soft" surface.
- Polyethylene is prone to shedding over time, when articulating against a coarse metal surface, causing debris and implant loosening.

The oxidizing process of zirconium creates a distinct set of characteristics that is not identical to those of other ceramics, particularly when the zirconium ceramic is created as a surface on a metal foundation. According to the manufacturer, the zirconium oxide forms an articulating surface that overcomes the disadvantages of metal and solid ceramic surfaces and is additionally biocompatible.

Patient Population:

Products in this category would have application limited to those patients whose life expectancy is presumably greater than the typical joint replacement patient. The exception to any age cohort would be patients who exhibit metal allergies that would likely result in early revision. Patients undergoing joint replacement with oxidized zirconium have the same risk profile as those who present for joint replacement with the current bearing surfaces when adjusted for known and

unknown metal allergies. It is arguable that co-morbidities may be less frequent in these patients due to a lower mean age at the time of implantation.

Several international registries including Australia separately track oxidized zirconium as a bearing surface.

Outcomes:

Studies in the literature support the use of oxidized zirconium articulation surfaces in hip arthroplasty and resurfacing. For example, a recent study showed that, at a minimum follow-up of 2 years, clinical outcomes for THA procedures comparing oxidized zirconium articulation surfaces and chromium-cobalt femoral heads that oxidized zirconium appears statistical superior.

Current Coding

Currently code 00.77, Hip bearing surface, ceramic-on-polyethylene, is assigned for hip replacement patients using the oxidized zirconium ceramic bearing surface.

Coding Options:

Option 1: Do not create a new code for oxidized zirconium ceramic bearing surface since it is appropriately grouped with other ceramic bearing surfaces. Continue to assign code 00.77, Hip bearing surface, ceramic-on-polyethylene, to capture this bearing surface. Add an inclusion term to clarify the use of this code as follows.

00.77 Hip bearing surface, ceramic-on-polyethylene
Add inclusion term Hip bearing surface, oxidized zirconium-on-polyethylene

Option 2: Create a new code to capture the oxidized zirconium ceramic bearing surface in hip replacements. Add excludes note under 00.77 to exclude this type of ceramic bearing surface.

New code 00.78 Hip bearing surface, oxidized zirconium
Oxidized zirconium ceramic on polyethylene

CMS Recommendation:

Option 1: Do not create a new code for this type of ceramic bearing surface. Continue assigning code 00.77 for oxidized zirconium ceramic hip bearing surface. Add an inclusion term to clarify this code assignment.

Interim Coding:

Continue to assign code 00.77 for oxidized zirconium hip bearing surface.



Insertion of Sling / Tape for Correction of Urinary Stress Incontinence

Issue: There is not a specific procedure code for males or females to capture the minimally invasive sling operation for urinary stress incontinence. A requestor recommends a new, specific ICD-9-CM code for Insertion of Sling / Tape for Correction of Urinary Stress Incontinence for both male and female, and a new specific ICD-9-CM code for Removal or Revision of sling or tape for both male and female.

New Technology Application:

No.

Food & Drug Administration (FDA) Approval: Not applicable.

Background: Stress incontinence in women often results from the bladder losing support and gradually dropping toward the vagina. In men this occurs in the setting of prior pelvic surgery including post-TURP (transurethral resection of the prostate) and post-radical prostatectomy. Traditional repair involves an open surgical procedure in which the bladder is stabilized by adjacent muscle or bone. Minimally invasive procedures involve the use of artificial mesh tapes or slings to stabilize the bladder and/or apply pressure to the urethra. Depending on the selected approach, the mesh is typically inserted through a pair of suprapubic incisions (women) or through small incisions in the perineum. The slings may be self-stabilizing or anchored by sutures, and may be compressive or non-compressive, again depending on the specific mesh and approach chosen by the surgeon. Since these procedures are minimally invasive, they are generally performed as outpatients with or without an overnight stay, but certain patients may require an inpatient setting.

Current Coding:

Female procedure codes: 59.79, Other repair of urinary stress incontinence
 70.95, Insertion of Synthetic Graft or Prosthesis.
Male procedure code: 59.79, Other repair of urinary stress incontinence

The following options were considered for sling operation for stress incontinence. These options could be used for either males or females:

Option 1: Do not create a new code and continue to existing code as described in Current Coding above.

Option 2: Create a new code under category under 58.9, Other operations on urethra and periurethral tissue, to capture the insertion of sling / tape for correction of urinary stress incontinence:

58.9 Other operations on urethra and periurethral tissue

New Code 58.94 Insertion of sling / tape for correction of urinary stress incontinence

Create a new code under category 59.9, Other operations on urinary system, to capture the removal or revision of sling / tape for correction of urinary stress incontinence.

59.9 Other operations on urinary system

New Code: 59.96 Removal or revision of sling / tape for correction of urinary stress incontinence

Option 3: Create a new code under category under 59.7 to capture the insertion of sling / tape for correction of urinary stress incontinence:

59.7 Other Repair of Urinary Stress Incontinence

New Code 59.73 Insertion of sling / tape for correction of urinary stress incontinence

Create a new code under category 59.9, Other operations on urinary system, to capture the removal or revision of sling / tape for correction of urinary stress incontinence.

59.9 Other operations on urinary system

New Code: 59.96 Removal or revision of sling / tape for correction of urinary stress incontinence

CMS Recommendation: Option 3 as above

Interim Coding: In the interim, continue to use the female procedure codes: 59.79, Other repair of urinary stress incontinence and 70.95, Insertion of synthetic graft or prosthesis and the male procedure code: 59.79, Other repair of urinary stress incontinence.

Sleeve Gastrectomy

Issue: There is not a specific ICD-9-CM procedure code for sleeve gastrectomy. A requestor recommends that a unique procedure code be created to identify these surgical procedures through both open and laparoscopic approaches.

New Technology Application? Not applicable

Food and Drug Administration Approval: Not applicable

Background: The sleeve gastrectomy and laparoscopic banding gastroplasty are common procedures used to treat obesity. Sleeve gastrectomy, also called vertical sleeve gastrectomy (VSG) can be performed using an open or laparoscopic approach. It is a surgical procedure where the left side of the stomach or the greater curvature is removed. The remaining portion of the stomach is approximately the size and shape of a banana. This operation is less complex than the gastric bypass or duodenal switch because ‘rerouting’ or reconnecting of the intestines is not performed. Unlike the laparoscopic banding procedure there is no implantation of an artificial device inside the abdomen in a sleeve gastrectomy.

The sleeve gastrectomy can be performed as a definitive (one-stage) procedure, or as the first part of a 2-stage operation, which is typically performed on obese patients with BMI of over 60%. In the 2-stage operation, the sleeve gastrectomy is performed first, allowing the patient to lose significant weight prior to undergoing the second procedure months later – a gastric bypass or duodenal switch.

Current Coding: Sleeve gastrectomy is currently captured under code 43.89, Other partial gastrectomy.

Coding Options:

Option 1: Do not create a new ICD-9-CM procedure code and continue to capture under code 43.89, Other partial gastrectomy.

Option 2: Create two new procedure codes for the laparoscopic and open approaches as follows:

43.8 Other partial gastrectomy

New code	43.82 Laparoscopic vertical (sleeve) gastrectomy
	Excludes: laparoscopic banding (44.95)
	laparoscopic gastric restrictive procedure (44.95)

New code 43.83 Other vertical (sleeve) gastrectomy
 Open vertical (sleeve) gastrectomy

 43.89 Other partial gastrectomy
 Partial gastrectomy with bypass gastrogastrostomy

Add exclusion term Excludes: laparoscopic sleeve gastrectomy (43.82)
 vertical sleeve gastrectomy (43.83)

CMS Recommendation: We recommend option 2; create two new procedure codes, as described above.

Interim Coding: In the interim, continue to use code 43.89, Other partial gastrectomy, for sleeve gastrectomy procedures.

Electromagnetic Navigation Bronchoscopy

Issue: Current ICD-9-CM procedure codes do not describe a new procedure in which a solitary pulmonary nodule (SPN) also referred to as a peripheral lung lesion, is visualized and accessed through the patients' airway, using an electromagnetic tip tracked series of instruments, while the patient continues to breathe normally.

New Technology Application: Yes. This may be a new technology and / or newly approved by the FDA, but it is not a New Tech Application for CMS IPPS FY2012.

Food & Drug Administration (FDA) Approval: Yes. The FDA defines the product codes associated with the Veran Medical Technologies Ig4 Spin Drive under Product Code of KTI which identifies their Electromagnetic (EM) Tip Tracked steerable devices as an accessory to a bronchoscope or as functioning as a bronchoscope.

Background of Technology and Procedure: Electromagnetic navigation is an aid to bronchoscopy that allows the bronchoscopist to rapidly reach lesions in the periphery of the lung. It involves the use of a sensor probe on an ultrathin bronchoscope. The patient is positioned on an electromagnetic grid that tracks the position of the probe. Computer software then superimposes that position on a previously acquired CT image of the lung. This is particularly valuable in accessing solitary pulmonary nodules that are smaller and located more peripherally than can be reached by conventional bronchoscopy, and providing an additional alternative to image-guided bronchoscopy using CT fluoroscopy and transthoracic approaches.

Current Coding: Per the requestor, the current ICD-9-CM procedure codes 33.22, Fiber-optic bronchoscopy, and 33.23, Other bronchoscopy, or rigid bronchoscopy, do not define the technique of the new procedure. The new device is flexible, steerable, and can always visually track the exact location of the small solitary pulmonary nodule while the interventional pulmonologist is advancing the device through the patient's trachea, bronchus and into the distal sub-segmental branch of the bronchial tree. The new technique allows an easier access through the patient's airway, to the central region and peripheral regions of the bronchial tree.

Coding Options:

Option 1. Do not create a new ICD-9-CM procedure code but continue to use existing ICD-9-CM procedure codes and their respective excludes notes to capture the electromagnetic tip tracked instrument 33.22 Fiber-optic bronchoscopy, or 33.27, Closed endoscopic biopsy of lung, to identify this technology.

Additional index or tabular changes can be made to assist with this code assignment.

	33.22 Fiber-optic bronchoscopy
Add inclusion term	<u>Electromagnetic tip tracked procedure</u>

Add inclusion term 33.24 Closed [endoscopic] biopsy of bronchus
Bronchoscopy (fiberoptic) (rigid) with:
electromagnetic navigation with biopsy

Revise inclusion term 33.27 Closed endoscopic biopsy of lung
Fiber-optic (flexible) bronchoscopy with:
Add inclusion term electromagnetic navigation with biopsy
Revise inclusion term indent fluoroscopic guidance with biopsy

Option 2. Create new subcategory and five new codes.

New subcategory 33.8 Electromagnetic tip tracked procedures

New code 33.80 Electromagnetic tip tracked procedure of the bronchus or lung

New code 33.81 Electromagnetic tip tracked biopsy of bronchus

New code 33.82 Electromagnetic tip tracked biopsy of lung

New code 33.83 Electromagnetic tip tracked excision or destruction of lesion or
tissue of bronchus

New code 33.84 Electromagnetic tip tracked excision or destruction of lesion or
tissue of lung

Option 3. Create a new code in subcategory 32.0, Local excision or destruction of lesion or tissue of bronchus.

32.0 Local excision or destruction of lesion or tissue of bronchus

New Code 32.07 Electromagnetic tip tracked excision or destruction of lesion or tissue of
bronchus

CMS Recommendation: Option 1; do not create a new code.

Interim Coding: Continue to use existing ICD-9-CM procedure codes and their respective excludes notes to capture the electromagnetic tip tracked instrument 33.22 Fiber-optic bronchoscopy, or 33.27 Closed Endoscopic Biopsy of Lung to identify this technology.

Ultrasound-enhanced Thrombolysis

Issue: Currently there is not a unique ICD-9-CM procedure code to specifically identify ultrasound-augmented thrombolysis. Should a new code be created to describe this technology?

New Technology Application? No.

FDA Approval: The Company developing this device has currently filed an IND and is preparing to run a PHASE III trial with the device described in this document as a follow-up to the Phase II CLOTBUST study. This study is not yet underway.

Background: Ultrasound energy can be applied to mechanically disrupt or “lyse” blood clots in the vasculature. When the acoustic energy from ultrasound is directed at the site of a blood clot, a radiation force is created that produces micro-streaming of the blood fluids at the site of the clot. It is believed that this effect enhances the penetration of endogenous or exogenous thrombolytics such as tissue plasminogen activator (tPA or alteplase) into the clot itself, thus accelerating the clot lysis process.

To administer this therapy, tPA is delivered intravenously as a continuous infusion while the ultrasound is administered externally through the skull, representing a non-invasive, easy-to-use therapy that can be rapidly employed by emergency room staff.

Ultrasonic Head frame

This head frame will employ multiple transducers to sequentially transmit therapeutic levels of transcranial ultrasound energy. The device will be comprised of multiple transducers fixed into three different arrays mounted on an adjustable head frame. The transducer arrays will be fixed in place to take advantage of known acoustic windows on the human head and will administer therapeutic ultrasound to the principal regions in which vessel occlusions of the brain are generally known to occur.

In general, the dose of ultrasound energy imparted on the wearer will be equivalent to the dose given by any currently marketed non-imaging approved transcranial Doppler TCD diagnostic ultrasound devices, or less.

The device will consist of a head frame with integrated transducers, battery, and a small pre-programmed microprocessor to control the unit. The device differs from conventional TCD systems in that it features an array of ultrasound transducers over each of the three “acoustic windows” of the skull, allowing ultrasound energy to be directed to the intracranial vascular territories which are associated with ischemic strokes where the NIH stroke score is >10, including right and left M1 and M2 segments of the middle cerebral artery and the posterior basilar circulation. The head frame sequentially applies ultrasound to these regions.

Method of Employment

It is expected that when a stroke patient presents to the emergency room in hospitals equipped with computed tomography (CT) or magnetic resonance imaging capability (MRI), the type of stroke (hemorrhagic or ischemic) can be confirmed and, if ischemic, the approximate location of the thrombus can be located or isolated to one of the three regions to be insonated by the ultrasonic head frame. If the stroke is ischemic, the ER team will administer the tPA (if eligible), place the head frame on the patient, press the “start” button, and the device will begin sequentially insonating from each transducer.

When the 120 minute insonation/treatment cycle has been completed, the head frame is removed, lightly cleaned, and placed on a base charging station for recharging.

Clinical Background

Tissue plasminogen activator has been shown to be an effective treatment in stroke when used within certain guidelines. In the pivotal NINDS-rt-PA Stroke Study, treatment with intravenous (IV) tPA resulted in a 11-13% absolute increase in favorable outcome after 3 months in comparison to the control group. The rate of (sICH) was found to be 6.4% in that study. Ultrasound (US)-enhanced thrombolysis is an approach to increasing the thrombolytic efficacy of tPA. The goal is to increase rates of recanalization and favorable clinical outcome without an increase in the symptomatic intracranial hemorrhage (sICH) rate.

Current Coding: ICD-9-CM procedure code 00.01, Therapeutic ultrasound of vessels of head and neck, is the code used to capture this procedure.

Coding Options:

Option 1: Do not create a new ICD-9-CM procedure code. Use code 00.01, Therapeutic ultrasound of vessels of head and neck to identify the ultrasound technology. Use code 99.10, Injection or infusion of thrombolytic agent, to describe the intravenous delivery of thrombolytics such as tissue plasminogen activator (tPA or altepase).

00.01 Therapeutic ultrasound of vessels of head and neck
 Anti-restenotic ultrasound
 Intravascular non-ablative ultrasound

Excludes:

Diagnostic ultrasound of:

 eye (95.13)
 head and neck (88.71)

That of inner ear (20.79)

Ultrasonic:

 angioplasty of non-coronary vessel (39.50)
 embolectomy (38.01, 38.02)
 endarterectomy (38.11, 38.12)
 thrombectomy (38.01, 38.02)

Option 2: Create a new ICD-9-CM procedure code describing the procedure of combining transcranial Doppler (TCD) ultrasound with intravenous administration of tPA resulting in a greater rate of recanalization of occluded vessels in ischemic stroke patients compared to patients that received tPA therapy alone (p=0.002).

99.7 Therapeutic apheresis or other injection, administration, or infusion of other therapeutic or prophylactic substance

New Code: 99.70 Intravenous delivery of thrombolytic agents with the use of therapeutic ultrasound

Add exclusion term: 99.10 Injection or infusion of thrombolytic agent
Excludes: intravenous delivery of thrombolytic agents with use of therapeutic ultrasound (99.70)

CMS Recommendation: Option 1, do not create a new code.

Interim coding: In the interim, continue to use code 00.01, Therapeutic ultrasound of vessels of head and neck, to capture this procedure. Code 99.10, Injection or infusion of thrombolytic agent, can be reported to capture any use of thrombolytics.

External Ventricular Drainage

Issue: We received a request to clarify the ICD-9-CM volume 3 procedure Index and Tabular to direct coders to the appropriate code assignment for procedures involving ventricular drainage by the use of a catheter or shunt. Currently, procedure code 02.2, Ventriculostomy, and code 02.39, Other operations to establish drainage of ventricle, describe these types of procedures. Should new codes be created and/or revisions made to the existing codes to further clarify correct code assignment?

New Technology application? No.

Food & Drug Administration (FDA) Approval? Not applicable.

Background: ICD-9 procedure codes evolved out of the common names for procedures in use at the time the taxonomy was created. Since the taxonomy was descriptive rather than prescriptive, inconsistencies, overlaps and connotations in contemporary usage were at times reflected in the taxonomy. One such instance occurred with ventricular drainage procedures and has led to ongoing questions regarding code assignment.

An -ostomy is a “cut” (tome from latin tomus) to create an ostium (“door”). As typically used, code 02.2, Ventriculostomy, referred to an open procedure that created a passage between the ventricle and another intracranial space, such as in a Third Ventriculostomy. Since a shunt is a procedure that creates a bypass pathway for a flowing fluid, these were creating intracranial shunts although this was not the preferred terminology. Code 02.2, Ventriculostomy, was contrasted with code category 02.3, Extracranial Ventricular Shunt, in which a tube is inserted into the ventricle and tunneled under the skin in order to shunt the CSF elsewhere in the body, such as the abdomen. Since a catheter is a tube used to “let out” or “let down” fluids, the shunt tube was a type of catheter but again was not referenced in this manner.

The other set of common procedures involved simply inserting tubes (catheterization) into the ventricle, usually through a burr hole, to infuse drugs, measure pressures, or drain CSF. These procedures were not characterized as ostomies nor have they appropriately been considered to be shunts. Catheterization codes, including 01.28, Placement of intracerebral catheter(s) via burr hole(s) and 01.26, Insertion of catheters into cranial cavity or tissue, have significant semantic overlap but were also limited by common usage.

Against this backdrop the question arises of correct coding for a ventricular drain, in which a catheter is passed through a small burr hole into the ventricle and fluid is allowed to drain into a container outside the body.

Coding Options:

Option 1. Do not create new codes or revise the existing codes. Continue to use existing codes 02.2, Ventriculostomy, and codes from subcategory 02.3, Extracranial Ventricular Shunt, to identify these drainage procedures.

Option 2. Create a new subcategory at 02.2 and create two new codes to distinguish between an external ventricular drain and an intracranial shunt. Revise the code title to existing code 02.39 to help clarify that this code is assigned for extracranial shunt procedures.

New subcategory	02.2 Ventriculostomy
Delete inclusion term	Anastomosis of ventricle to:
Delete inclusion term	cervical subarachnoid space
Delete inclusion term	cistern magna
Delete inclusion term	Insertion of Holter Valve
Delete inclusion term	Ventriculocisternal intubation
New code	02.21 Insertion or replacement of external ventricular drain [EVD] External ventricular drainage [EVD] setup Replacement of external ventricular drain Ventricular catheter placement for: drainage of cerebrospinal fluid [CSF] injection of medication or other substance sampling of cerebrospinal fluid [CSF]
New code	02.22 Intracranial ventricular shunt or anastomosis Anastomosis of ventricle to: cervical subarachnoid space cistern magna Insertion of Holter valve into intracranial system Shunt between two intracranial ventricles That by endoscopy Third ventriculostomy Ventriculocisternostomy
	02.3 Extracranial ventricular shunt Includes: That with insertion of valve
Revise code title	02.39 Other operations to establish drainage of ventricle <u>Ventricular shunt to extracranial site NEC</u> Ventricle to bone marrow shunt
Delete inclusion term	Ventricular shunt to extracranial site NEC

Revise code 02.42 by deleting an existing inclusion term that may cause confusion.

02.4 Revision, removal, and irrigation of ventricular shunt
Excludes: revision of distal catheter of ventricular shunt (54.95)

Delete inclusion term

02.42 Replacement of ventricular shunt
Reinsertion of Holter valve
~~Replacement of ventricular catheter~~
Revision of ventriculoperitoneal shunt at
ventricular site

CMS Recommendation: Option 2 as shown above.

Interim coding: In the interim, continue to use existing codes 02.2, Ventriculostomy, and a code from subcategory 02.3, Extracranial ventricular shunt, to identify these drainage procedures.

Embolization of Uterine Artery

Issue: There is not a unique ICD-9-CM procedure code to identify uterine artery embolization. Should a new code be created to describe this procedure?

New Technology Application? No

FDA Approval? Not applicable

Current coding: There is confusion among coders on correct code assignment for uterine artery embolization procedures. This issue has been addressed in AHA's *Coding Clinic for ICD-9-CM* however; comments continue to be received regarding the advice provided to assign code 99.29, Injection or infusion of other therapeutic or prophylactic substance, versus code 39.79, Other endovascular procedures on other vessels. As neither one of these existing procedure codes clearly describes that a uterine artery embolization was performed, should a new code be created to uniquely identify this procedure?

Coding options:

Option 1. Do not create a new code. Continue to use existing code 99.29, Injection or infusion of other therapeutic or prophylactic substance, to identify uterine artery embolization.

Option 2. Create a new code to specifically identify uterine artery embolization was performed. The new code could be placed in the Operations on the Female Genital Organs section of the code book.

New code	68.24 Uterine artery embolization [UAE] Includes that with or without coils
Add exclusion term	39.79 Other endovascular procedures on other vessels Excludes: <u>uterine artery embolization (68.24)</u>
Add exclusion term	99.29 Injection or infusion of other therapeutic or prophylactic substance Excludes: <u>uterine artery embolization (68.24)</u>

Option 3. Create two new codes that specifically identify uterine artery embolization with and without coils. Add exclusion terms to codes 39.79 and 99.29.

New code	68.24 Uterine artery embolization [UAE] with coils
Add exclusion term	39.79 Other endovascular procedures on other vessels Excludes: <u>uterine artery embolization with coils (68.24)</u>

New code	68.25 Uterine artery embolization [UAE] without coils
	99.29 Injection or infusion of other therapeutic or prophylactic substance
Add exclusion term	Excludes: <u>uterine artery embolization without coils (68.25)</u>

CMS Recommendation: Option 3, create two new codes as stated above.

Interim coding advice: Continue to use code 99.29, Injection or infusion of other therapeutic or prophylactic substance, for uterine artery embolization procedures by injection of a substance. Assign code 39.79, Other endovascular procedures on other vessels, for any uterine artery embolization procedures using coils.

Open Left Atrial Appendage Occlusion with “U” Fastener Implant

Issue: The occlusion, isolation or removal of the left atrial appendage (LAA) has been used as an alternative to oral anticoagulation therapy for patients with atrial fibrillation, and is standard practice in the surgical maze procedure. The FDA has approved staple-based and linear pressure clip devices, but there is no unique code which recognizes a silicone “U” fastener with connectors. Should a unique procedure code be created to identify this device?

New Technology Application? No

Food & Drug Administration (FDA) Approval: The TigerPaw® System technology was cleared for marketing by the FDA on October 29, 2010. The device may currently be in clinical trials.

Background: Atrial fibrillation (AF) is the most common type of arrhythmia, defined as an abnormality with the rate or rhythm of the heartbeat. In people with AF, it is common for blood clots to form in the left atrial appendage, a small sac attached to the left atrium. These clots can dislodge and travel directly to the brain where they obstruct blood flow, thus causing an ischemic stroke. Currently, long term oral anticoagulation (OAC) treatment is the most effective means of protection in patients with AF at high risk of stroke. However, as many as 20% of patients with AF are contraindicated, ineligible, or intolerant to OAC therapy.

The LAAX, Inc TigerPaw® System technology consists of a delivery tool and an implantable silicone fastener. The fastener consists of a series of evenly spaced individual connectors embedded in a silicone housing which has a “U” shaped connector at one end of the delivery jaws.

Patients having exclusion of the LAA via this device would be undergoing a concomitant open cardiac surgical procedure. Direct visualization by the surgeon is required.

Coding Options:

Option 1: Do not create a new procedure code.

Option 2:

Use existing ICD-9-CM procedure code 37.36, Excision or destruction of left atrial appendage (LAA), with revisions.

37.3	Pericardiectomy and excision of lesion of heart
Revise title	37.36 Excision, or destruction <u>or occlusion</u> of left atrial appendage (LAA)
Add note:	<u>Note: includes procedures done concomitantly with other cardiovascular procedures, or stand-alone LAA procedures</u>
Add includes note	<u>That using a silicone “U” fastener</u>

Revise code also note Code also any concomitant procedure performed, fluoroscopy (87.49) or transesophageal echocardiography (TEE) (88.72)

CMS Recommendation: Option 2, as listed above.

Interim Coding: Use existing procedure code 37.36, Excision or destruction of left atrial appendage (LAA).

Percutaneous Left Atrial Appendage (LAA) Exclusion with Femoral and Epicardial Access

Issue: There is currently no specific code that uniquely identifies a stand-alone procedure for percutaneous left atrial appendage (LAA) exclusion using epicardial and femoral access. The existing procedure code was intended for LAA exclusion in conjunction with another (open) cardiac procedure.

New Technology Application? No.

Background: Atrial fibrillation (AF) is the most common type of arrhythmia, defined as an abnormality with the rate or rhythm of the heartbeat. In people with AF, it is common for blood clots to form in the left atrial appendage, a small sac attached to the left atrium. These clots can dislodge and travel directly to the brain where they obstruct blood flow, thus causing an ischemic stroke. Currently, long term oral anticoagulation (OAC) treatment is the most effective means of protection in patients with AF at high risk of stroke. As many as 20% of patients with AF are contraindicated, ineligible, or intolerant to OAC therapy.

The other option for contraindicated patients is LAA exclusion at the time of a concomitant open sternotomy surgical procedure. However, many patients who are ineligible for long-term OAC and at high risk of stroke do not require a cardiac surgical procedure beyond LAA exclusion to protect them from a potentially fatal thrombus escaping the LAA. Percutaneous suture exclusion of the LAA is technology new option.

Following femoral access, a guide wire with a small magnet is placed in the LAA through the transseptal access and an occlusion balloon is advanced over the guide wire to the LAA. Using TEE, the pericardial access site is made and a guide cannula is placed into the pericardial sac. Another magnetic guide wire provides the delivery of the suture delivery device to the base (opening) of the LAA. Once in position, the suture snare is closed. A drainage catheter through the pericardial access is inserted into the pericardium for monitoring of pericardial or pleural effusion.

Current Coding:

If the ICD-9-CM procedure Index is followed, the coding instructions are clear: Excision, appendage, left atrial = 37.36. There is some reluctance to use this code as it has been felt to not describe a stand-alone OR procedure and does not specify occlusion of LAA.

Coding Options:

Option 1: Do not create a new code or modify existing codes. Continue to use code 37.36 as specified by the Index.

Option 2:

Revise existing code 37.36 to reflect the occlusion of the left atrial appendage via percutaneous approach.

37.3 Pericardiectomy and excision of lesion of heart
Revise title 37.36 Excision, ~~or~~ destruction or occlusion of left atrial appendage (LAA)
Add note: Note: includes procedures done concomitantly with other cardiovascular procedures, or stand-alone LAA procedures
Add includes note That by percutaneous and/or endovascular approach
Revise code also note Code also any concomitant procedure performed, fluoroscopy (87.49) or transesophageal echocardiography (TEE) (88.72)

Option 3: Make no major changes to code 37.36. Specify that it is a code that reflects a procedure done thoroscopically or as an open procedure, including concomitantly with another open procedure. Create a new code at 37.38 identifying the stand-alone LAA exclusion.

37.3 Pericardiectomy and excision of lesion of heart
Revise title 37.36 Open or thoracoscopic ~~E~~xcision, obstruction or destruction of left atrial appendage (LAA)
New code 37.38 Percutaneous exclusion of left atrial appendage (LAA)
That by femoral and/or epicardial access
Code also any fluoroscopy (87.49) or transesophageal echocardiography (TEE) (88.72)

CMS Recommendation:

CMS recommends Option 2, as described above. Revision of the existing code 37.36, Excision or destruction of left atrial appendage (LAA) appropriately describes the procedure performed.

Interim Coding:

Use existing code 37.36, as specified by the ICD-9-CM procedure Index.

Ultrasonic Wound Debridement

Issue: Misonix Ultrasonic Wound Debridement System utilizes an ultrasonic metallic probe and a water irrigation system to debride wounds. This procedure is currently captured by code 86.28, Nonexcisional debridement of wound, infection, or burn. The manufacturer has requested that indexing and tabular entries be changed so that this procedure would be captured by code 86.22, Excisional debridement of wound, infection, or burn instead of code 86.28.

New Technology Application? No

Food & Drug Administration (FDA) approval: Yes, the Misonix SonicOne contact wound debridement system has been cleared under 510K #: K050776.

Background: Contact ultrasonic wound debridement systems use ultrasonically energized instruments and utilize water irrigation for tissue excision. The ultrasonic probe is brought into direct contact with the tissue targeted for debridement. The probe's vibratory motion acts as a micro jackhammer, creating local cavitation and hydrodynamic effects that dissect tissue at the point of contact. The manufacturer states that the system is applicable for tissue excision outside the wound margin, as well as within it. However the manufacturer also emphasizes that an advantage of this system is that "viable tissue structures are more resistant to destruction" and the system "flushes the wound of fibrin deposits...while preserving the granulation tissue." Advice in both the *American Hospitals Association's Coding Clinic for ICD-9-CM*, as well as indexing and tabular entries in ICD-9-CM for water scalpel (jet) procedures, leads to a code assignment of 86.28. This Ultrasonic Wound Debridement System is also captured by code 86.28. The manufacturer disagrees with this code assignment and believes that this technology performs an excisional debridement.

Coding options:

Option 1.

Continue capturing Misonix Ultrasonic Wound Debridement System and other ultrasonic mechanical debridement systems through code 86.28, Nonexcisional debridement of wound, infection, or burn.

Add an index entry and a tabular entry under code 86.28 for Ultrasonic mechanical debridement to further clarify this code assignment. As indicated above, the process is not typically used to excise block elements that can be available for examination but to pulverize (ablate) devitalized tissue and debris into small particles for removal by the irrigation fluid. Additionally, although this product does provide the capability of creating incisions (cuts) into vital tissue which is the hallmark of excisional debridement, it is not designed, marketed or typically used to leverage that ability.

Option 2.

Change the indexing and tabular entries so that ultrasonic mechanical debridement systems would be assigned to code 86.22, Excisional debridement of wound, infection, or burn instead of 86.28, Nonexcisional debridement of wound, infection, or burn.

CMS Recommendation: Option 1, Continue capturing Misonix Ultrasonic Wound Debridement System and other ultrasonic mechanical debridement systems through code 86.28, Nonexcisional debridement of wound, infection, or burn. Add an index entry and a tabular entry under code 86.28 for Ultrasonic mechanical debridement.

Interim coding: Capture the use of Ultrasonic Wound debridement through code 86.28, Nonexcisional debridement of wound, infection, or burn.

Hydrosurgery/Versajet Debridement

Issue: Versajet™ Hydrosurgery System, is a high-pressure, high velocity water jet that assists in the removal of devitalized tissue in the treatment of certain wounds. The System uses pressurized streams of sterile fluid to ablate and remove tissue and foreign matter from wounds and to remove material in a variety of surgical applications. The system depends on the use of repeated passes of the hydrostatic handpiece over the surface of the wound in order to remove surface elements in the manner of shaving or curettage. This procedure is currently captured in code 86.28, Nonexcisional debridement of wound, infection, or burn.

New Technology? No.

Food & Drug Administration (FDA) approval: The Versajet received 510(k) clearance on June 22, 2001.

Background: Advice in both the *American Hospitals Association's Coding Clinic for ICD-9-CM*, as well as indexing and tabular entries in ICD-9-CM for water scalpel (jet) procedures, leads to a code assignment of 86.28 as previously stated. The manufacturer disagrees with this code assignment and believes that the VersaJet performs an excisional debridement. The manufacturer states that other water jet procedures will not remove tissue outside the wound margins and thus do not meet the definition of excisional debridement. However, they believe that VersaJet differs from other water jet devices in that it does cut outside the margins of the wound and meets the definition of "excisional" debridement as defined in the guidance. In order to more clearly capture the debridement performed through the VersaJet Hydrosurgery System, they have requested the creation of a new code.

The manufacturer states that VersaJet Hydrosurgery System uses pressurized streams of sterile fluid to "cut, ablate and remove" tissue and foreign matter from wounds and "resects and removes" material in a variety of surgical applications. The manufacturer states that the device provides cutting, irrigation and evacuation in the same tool. The stream of fluid simultaneously washes the tissue surface and removes foreign material from the wound or surgical site. The stream of saline simultaneously washes the tissue surface and vacuums away foreign material, including contamination and infected and necrotic tissue from the wound. The manufacturer states that the fluid acts to ablate the surface of the tissue and propel the tissue and debris out of the surgical site. The debris and fluid are directed immediately within the instrument into a flexible tube, which carries the effluent to the drain or a collection canister.

The System employs two basic system components: the reusable power console unit and the sterile, disposable pump cartridge, handpiece and tubing assembly.

Coding Options:

Option 1. Do not create a new code. Continue capturing the use of hydrosurgery for debridement, including that by Versajet, under code 86.28.

Option 2. Create a new code for debridement performed using hydrosurgery as follows:

New Code: 86.20 Debridement of wound, infection or burn using hydrosurgery

CMS recommendation: Option 1. Do not create a new code. The process does not excise block elements that can be available for examination but pulverizes (ablates) the tissue into small particles for vacuum removal by Venturi effect. Additionally, the process is approved for the removal of devitalized tissue and debris and is marketed for its ability to spare vital tissue. It does not provide the capability of creating incisions (cuts) into vital tissue that are both the hallmark and risk of excisional debridement. Continue capturing the use of hydrosurgery for debridement, including that by Versajet, under code 86.28.

Interim coding: Capture the use of hydrosurgery for debridement, including that by Versajet, under code 86.28, Nonexcisional debridement of wound, infection, or burn

Non-excisional Debridement

Issue: Currently, procedure code 86.28, Nonexcisional debridement of wound, infection, or burn, is located within subcategory 86.2, Excision or destruction of lesion or tissue of skin and subcutaneous tissue. A requestor believes it is inappropriate to report debridement of tissue other than skin and subcutaneous tissue with a code located in the chapter for Operations on the Integumentary System. The request is to create new codes to report non excisional debridement of the deeper tissue layers (soft tissue, fascia, muscle and bone). Should new codes be created?

Coding options:

Option 1. Do not create new codes. Continue to use existing code 86.28, Nonexcisional debridement of wound, infection, or burn, to identify all nonexcisional debridements.

Option 2. Create a new subcategory and codes to identify nonexcisional debridement of tendon, muscle, deep fascia and bone as outlined below. Add exclusion notes to code 86.28.

New subcategory	17.8 Nonexcisional debridement of wound, infection, or burn of soft tissue, fascia, muscle, or bone Debridement NOS Removal of devitalized tissue, necrosis, and slough by such methods as: brushing irrigation (under pressure) scrubbing washing Water scalpel (jet)
New code	17.81 Non excisional debridement of soft tissue Tendon
New code	17.82 Non excisional debridement of deep fascia
New code	17.83 Non excisional debridement of muscle
New code	17.84 Non excisional debridement of bone

CMS Recommendation: Do not create new codes. Debridement of the deeper tissues through a wound, infection or burn is part of the care of pathology in the integumentary system and is appropriately coded by 86.28.

Interim coding advice: Continue to use existing code 86.28, Nonexcisional debridement of wound, infection, or burn, to identify all nonexcisional debridements.

Cerebral and Somatic Oximetry

Issue: We received a request to create a unique ICD-9-CM procedure code for tissue oxygen saturation monitoring using Near-infrared spectroscopy. Near-infrared spectroscopy (NIRS) is currently captured in code 38.23, Intravascular spectroscopy, which has a table notation “Includes near infrared (NIR) spectroscopy.” NIR spectroscopy is used as a noninvasive method for the continuous, real-time monitoring of both cerebral and somatic tissue oxygenation in order to detect regional hypoperfusion.

We received a request to create new ICD-9-CM codes for Tissue Oxygen Saturation Monitoring using Near-Infrared Spectroscopy and Cerebral Oxygen Saturation Monitoring using Near-Infrared Spectroscopy. The requestor states that the creation of a new ICD-9-CM procedure code would allow reporting and tracking of the use of these procedures.

New Technology Application? No.

Food and Drug Administration (FDA) Approval: The Somanetics INVOS® Cerebral/Somatic Oximeter first received FDA 510(k) clearance in May of 1996; in 2008 the somatic oximeter was cleared for monitoring of oxygen saturation in tissues beneath the sensor in individuals at risk for reduced-flow or no-flow ischemic states.

Background: NIRS is a noninvasive, optical method for continuous, real-time monitoring of tissue oxygenation. The NIRS technique is based on the transmission and absorption of near infrared (NIR) light (700 – 950 nm) as it passes through tissue. Oxygenated and unoxygenated hemoglobin have different absorption characteristics, allowing the system to reflect the oxygen saturation of the target tissue. Because near infrared light penetrates tissue relatively easily, the sensor reflects the saturation of deep tissues rather than surface capillary beds. Variation in penetration does not permit the device to report absolute levels; rather NIRS is used to track trends in oxygenation levels once a baseline has been established at a particular site in a given patient.

NIRS was first used to perform cerebral oximetry in cardiac surgery arena, but has subsequently been used to monitor the adequacy of tissue perfusion on somatic sites such as muscle, kidneys and the intestines, and to provide an early alert to unexpected hemodynamic changes.

Coding Options

Option 1: Do not create new coding structures for Tissue Oxygen Saturation Monitoring Procedures using Near-Infrared Spectroscopy (NIRS). Continue capturing this through code 38.23, Intravascular spectroscopy.

Option 2: Create a new procedure code as follows:

New code XX.XY Tissue oxygen saturation monitoring using Near-Infrared Spectroscopy, where the terminal digit, Y, is used to specify the target organ.

Option 3: Create new procedure codes within each system, e.g. 01.NN Cerebral Oxygen Saturation Monitoring using Near-Infrared Spectroscopy

CMS Recommendation: Option 1. Do not create a new code.

Interim coding advice: Continue to use code 38.23, Intravascular spectroscopy.

Ultrasound Assisted Lysis of Intravascular Thrombus

Issue: There is currently no specific code which uniquely identifies the procedure of placement and monitoring of the ultrasound assisted lysis catheter for arterial and venous thrombus.

New Technology Application? No.

Background: Traditional catheter-directed thrombolysis is a minimally invasive endovascular procedure that dissolves blood clots in blood vessels to help improve blood flow and prevent damage to tissues and organs. In a catheter-directed thrombolysis procedure, x-ray imaging is used to help guide a catheter to the site of blood clots to deliver medication to dissolve the blockage.

A new FDA approved catheter based system combines a drug delivery catheter with an ultrasound core wire to augment drug-induced thrombolysis. The catheter has two components: 1) a drug delivery catheter and 2) an ultrasound core wire with small ultrasound transducers that emit high frequency, low power sound waves into the clot. Ultrasound makes the clot more porous and therefore able to absorb more lytic than clots that are not exposed to sound waves. *Percutaneous Mechanical Thrombectomy* (PMT) uses other devices to break up the thrombus. Ultrasound is not viewed by the manufacturer as ablative but as an adjunct to increase the efficacy of the pharmaceutical lytic.

Ultrasound Assisted Lysis of Arterial or Venous Thrombus Procedure Description

The physician accesses the femoral artery or vein and performs an angiogram to identify the anatomy and thrombus formation. A guidewire is placed through the access point and advanced through the thrombus. The delivery catheter is placed over the guidewire into the thrombus. Placement of the catheter is checked with fluoroscopy. The catheter is connected to the infusion pump and lytic and ultrasound therapy is started.

The patient is returned to the cath lab for additional angiograms to monitor progress. If the thrombus is still present after 24 hours of therapy, the clot is not receptive to CDT and another option may be considered. Once the thrombus is dissolved and any obstruction treated, the ultrasound assisted lysis catheter is removed from the patient.

Coding options

Option 1:

Do not create a new ICD-9-CM procedure code. Use codes in Category 00.0, Therapeutic ultrasound, (Codes 00.01 – 00.09) to identify the ultrasound technology.

Then use code 99.10, Injection or infusion of thrombolytic agent, to describe the intravenous delivery of thrombolytics such as tissue plasminogen activator (tPA or alteplase).

Addenda Item

At the September 15, 2010 C&M Meeting, we discussed Endovascular Embolization with Head or Neck Vessel Reconstruction – the Pipeline™ Embolization Device. This device mimics devices and stents that are already on the market, and we had difficulty during the discussion of differentiating this device from stents that are available for use in head or neck vessels. At the time of the meeting, CMS did not have a clear recommendation regarding creation of a new code or codes, and asked for input from the audience. The written responses we received were not overwhelmingly in support of creation of a new code. Additionally, the Company has not sought New Technology status from CMS for the 2012 fiscal year. At the time, the suggestion was brought to the floor to add a new procedure code as follows:

New code	39.77 Endovascular embolization with head or neck vessel vascular remodeling support Embolization stent Stent-like device That for repair of aneurysm
----------	---

CMS' Recommendation is Option 1; do not create a new procedure code to describe this device.

CMS' interim coding advice remains the same: continue to use procedure code 39.72, Endovascular embolization or occlusion of head and neck vessels, to describe this procedure. We are interested in hearing additional comments about this topic.

Proposed Addenda

Tabular

Revise code title	00.61 Percutaneous angioplasty or atherectomy of precerebral (extracranial) vessel(s)
Revise code title	00.62 Percutaneous angioplasty or atherectomy of intracranial <u>intracerebral</u> vessel(s)
Revise code also note	Code also any: Percutaneous insertion of intracranial <u>intracerebral</u> stent(s) (00.65)
Revise code title	00.64 Percutaneous insertion of other precerebral (extracranial) artery stent(s)
Revise code title	00.65 Percutaneous insertion of intracranial <u>intracerebral</u> artery stent(s)
Revise code also note	Code also any: Percutaneous angioplasty or atherectomy of intracranial <u>intracerebral</u> vessel(s) (00.62)
Add inclusion term	00.94 Intra-operative neurophysiologic monitoring
Add inclusion term	<u>That by:</u>
Add inclusion term	<u>brainstem auditory evoked potentials [BAEP]</u>
Add inclusion term	<u>electroencephalogram [EEG]</u>
Add inclusion term	<u>electromyogram [EMG]</u>
Add inclusion term	<u>motor evoked potentials [MEP]</u>
Add inclusion term	<u>nerve conduction study</u>
Add inclusion term	<u>somatosensory evoked potentials [SSEP]</u>
Add inclusion term	<u>transcranial Doppler</u>
Revise code title	13.65 Excision of secondary membrane [after cataract]
title	Capsulectomy <u>Capsulectomy</u> Folio issue – inclusion term was in
Add exclusion term	88.7 Diagnostic ultrasound Excludes: <u>that for intraoperative monitoring 00.94</u>



Add exclusion term 89.19 Video and radio-telemetered electroencephalographic monitoring
Excludes: intraoperative monitoring 00.94

Add exclusion term 93.08 Electromyography
Excludes: that for intraoperative monitoring 00.94

Index

Angioplasty (laser) – *see also* Repair, blood vessel
percutaneous transluminal (balloon)
Revise subterm basilar ~~00.61~~
Add subterm precerebral 00.61
Add subterm intracerebral 00.62

Add term Cleft lift 86.21

Add subterm Dopplergram, Doppler flow mapping – *see also* Ultrasonography
intraoperative transcranial 00.94

Add subterm Electroencephalogram (EEG) 89.14
monitoring (radiographic) (video) 89.19
intraoperative 00.94

Add subterm Infusion
lymphocyte 99.09

Add subterm Monitoring
electroencephalographic 89.19
intraoperative 00.94
Revise subterm ~~intraoperative anesthetic effect monitoring and titration 00.94~~
~~[89.14]~~
Revise subterm intraoperative
Revise subterm anesthetic effect monitoring and titration (IAEMT) 00.94
~~[89.14]~~
Add subterm neurophysiologic (BAEP) (brainstem auditory evoked potentials) (EEG) (electroencephalogram) (electromyogram) (EMG) (MEP) (motor evoked potentials) (nerve conduction study) (somatosensory evoked potentials) (SSEP) (transcranial Doppler) 00.94

Partial Code Freeze for ICD-9-CM and ICD-10 Finalized

The ICD-9-CM Coordination and Maintenance Committee will implement a partial freeze of the ICD-9-CM and ICD-10 (ICD-10-CM and ICD-10-PCS) codes prior to the implementation of ICD-10 on October 1, 2013. There was considerable support for this partial freeze. The partial freeze will be implemented as follows:

- The last regular, annual updates to both ICD-9-CM and ICD-10 code sets will be made on October 1, 2011.
- On October 1, 2012, there will be only limited code updates to both the ICD-9-CM and ICD-10 code sets to capture new technologies and diseases as required by section 503(a) of Pub. L. 108-173.
- On October 1, 2013, there will be only limited code updates to ICD-10 code sets to capture new technologies and diagnoses as required by section 503(a) of Pub. L. 108-173. There will be no updates to ICD-9-CM, as it will no longer be used for reporting.
- On October 1, 2014, regular updates to ICD-10 will begin.

The ICD-9-CM Coordination and Maintenance Committee will continue to meet twice a year during the partial freeze. At these meetings, the public will be asked to comment on whether or not requests for new diagnosis or procedure codes should be created based on the criteria of the need to capture a new technology or disease. Any code requests that do not meet the criteria will be evaluated for implementation within ICD-10 on and after October 1, 2014 once the partial freeze has ended.

Codes discussed at the September 15 – 16, 2010 and March 9 – 10, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be considered for implementation on October 1, 2011, the last regular updates for ICD-9-CM and ICD-10. Code requests discussed at the September 14 – 15, 2011 and additional meetings during the freeze will be evaluated for either the limited updates to capture new technologies and diseases during the freeze period or for implementation to ICD-10 on October 1, 2014. The public will be actively involved in evaluating the merits of any such requests during the period of the partial freeze.

ICD-10-CM & ICD-10-PCS Abbreviated Titles

60 character abbreviated titles have been developed for the 2011 version of both ICD-10-CM and ICD-10-PCS codes. The ICD-10 abbreviated titles for both ICD-10-CM and ICD-10-PCS have been posted on the CMS website at

http://www.cms.gov/ICD10/11b1_2011_ICD10CM_and_GEMs.asp and

http://www.cms.gov/ICD10/11b_2011_ICD10PCS.asp

CMS will maintain and update these abbreviated titles each year. Users should feel free to download and use these, and other ICD-10 files posted on the CMS website.

Version 28.0 ICD-10 MS-DRGs

The ICD-10 MS-DRGs v28 Definitions Manual (based on FY2011 MS-DRGs) is now posted on the Centers for Medicare & Medicaid Services (CMS) website at

http://www.cms.gov/ICD10/17_ICD10_MS_DRG_Conversion_Project.asp in the “Related Links Inside CMS” section. This update is part of the ICD-10 MS-DRG Conversion Project. In the Conversion Project, CMS is using the General Equivalence Mappings (GEMs) to convert CMS payment systems. CMS is sharing information learned from this project with other organizations facing similar conversion projects. CMS has also posted the ICD-10 FY 2011 Medicare Code Editor. Please note that the ICD-10 MS-DRGs will be subject to formal rulemaking.

The public is encouraged to review these files. A detailed presentation on this conversion activity will take place at the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee Meeting.

2011 GEMs Update

The updated 2011 General Equivalence Mappings (GEMs) are posted for review and public comment. The updated files contain all changes to date in response to public comment as mandated by the Affordable Care Act, for the period ending November 11, 2010, as well as additional ongoing internal review for accuracy and completeness.

Extensive public comment was received on the subject of the GEMs. Several organizations sent general letters of support for the GEMs and asked for ongoing maintenance and increased stakeholder input in improving and updating the GEMs. Providers, payers, vendors, independent consultants and other individuals in the healthcare community submitted comments and suggestions for improving the GEMs. Approximately 5,200 GEMs entries were the subject of public comment. Roughly 250,000 total entries comprise the GEMs.

The public support and collaboration on the GEMs led to many improvements in the accuracy and completeness of the GEMs as an ICD-10 transition resource. All comments and suggestions were reviewed and considered. Any recommended change to an entry that met GEMs inclusion criteria was incorporated into the updated 2011 files. Of the 5,200 comments submitted, roughly one third of the recommended changes were either implemented for this update or had been previously implemented in the September update, based on public comment received during the

past year. Approximately 850 recommended changes, or 16% of all comments received, were new changes implemented for the 2011 GEMs, and approximately 900 additional recommended changes, or 17% of all comments received, supported previous changes in the most recent updated GEMs files (posted on the CMS website in September, 2010).

The 2011 ICD-10 GEM files for both ICD-10-CM and ICD-10-PCS have been posted on the CMS website at http://www.cms.gov/ICD10/11b_2011_ICD10PCS.asp and http://www.cms.gov/ICD10/11b1_2011_ICD10CM_and_GEMs.asp

ICD-10-PCS Device Key Being Developed

A PCS device key similar in organization and function to the PCS Body Part Key is under development. The Device Key will help users choose the correct PCS device value for a given clinical or industry device name. The resulting entries will be a public domain reference to accompany the ICD-10-PCS tables. Entries will be available for lookup by both device value and clinical or industry device name. Similar in function to “includes” notes in ICD-10-CM, the ICD-10-PCS device key will be an official part of the classification system.

The draft device key will be posted with the FY2011 ICD-10-PCS update for public review and comment prior to the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee meeting. Industry input will be solicited on this draft document. A detailed presentation on the ICD-10-PCS device key and related reference material will take place at the September 14-15, 2011 ICD-9-CM Coordination and Maintenance Committee Meeting.

New FY2012 ICD-10-PCS Update Schedule

The ICD-10-PCS annual update for FY2012 will follow the same update schedule as the annual ICD-9-CM update for procedure codes. CMS is adopting this new ICD-10-PCS update schedule in preparation for the move to ICD-10 on October 1, 2013. ICD-10-PCS changes discussed at the September 15 – 16, 2010 and March 9 – 10, 2011 ICD-9-CM Coordination and Maintenance Committee meeting will be considered for inclusion in the FY 2012 ICD-10-PCS updates. The updated timeline will be as follows:

June 2011 The FY 2012 ICD-10-PCS final addenda will be posted on the ICD-10 CMS website at <http://www.cms.gov/ICD10/>

October 2011 The FY 2012 ICD-10-PCS GEMs and the Procedure Reimbursement Mappings will be posted on the ICD-10 CMS website at <http://www.cms.gov/ICD10/>

This new schedule of posting ICD-10-PCS final addenda in June and the updated ICD-10-PCS GEMs and Reimbursement mappings in October will continue each year up to and after the implementation of ICD-10.

Ankle, Hip and Knee Joint Replacement: Request for New Values in PCS Tables

The original proposal is to create codes that differentiate between cemented and uncemented metal on polyethylene ankle replacement procedures. After consultation with the orthopedic community, it was proposed that this information be included for hip and knee replacement procedures as well. According to the original proposal, the current industry standard is to use only metal on polyethylene implants for the ankle joint. Therefore, additional information regarding the synthetic ankle joint does not need to be added to the PCS tables at this time.

Option 1: Do not create new PCS device values to differentiate between cemented and uncemented synthetic joint replacement procedures.

Section: 0 Medical and Surgical

Body System: S Lower Joints

Operation: R Replacement: Putting in or on biological or synthetic material that physically takes the place and /or function of all or a portion of a body part

Body Part	Approach	Device	Qualifier
C Knee Joint, Right D Knee Joint, Left F Ankle Joint, Right G Ankle Joint, Left T Knee Joint, Femoral Surface, Right U Knee Joint, Femoral Surface, Left V Knee Joint, Tibial Surface, Right W Knee Joint, Tibial Surface, Left	0 Open	J Synthetic Substitute	Z No Qualifier

Body Part	Approach	Device	Qualifier
9 Hip Joint, Right B Hip Joint, Left	0 Open	J Synthetic Substitute	5 Metal on Polyethylene 6 Metal on Metal 7 Ceramic on Ceramic 8 Ceramic on Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
A Hip Joint, Acetabular Surface, Right E Hip Joint, Acetabular Surface, Left	0 Open	J Synthetic Substitute	F Metal G Ceramic H Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
R Hip Joint, Femoral Surface, Right S Hip Joint, Femoral Surface, Left	0 Open	J Synthetic Substitute	F Metal G Ceramic Z No Qualifier

Option 2: Add new device value to specify cemented implants for hip, knee and ankle joint body parts in table 0SR (32 codes). Specify uncemented implants using qualifier J Synthetic Substitute. This assumes that all implants are either cemented or uncemented.

Section: 0 Medical and Surgical

Body System: S Lower Joints

Operation: R Replacement: Putting in or on biological or synthetic material that physically takes the place and /or function of all or a portion of a body part

Body Part	Approach	Device	Qualifier
C Knee Joint, Right D Knee Joint, Left F Ankle Joint, Right G Ankle Joint, Left T Knee Joint, Femoral Surface, Right U Knee Joint, Femoral Surface, Left V Knee Joint, Tibial Surface, Right W Knee Joint, Tibial Surface, Left	0 Open	<u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	Z No Qualifier

Body Part	Approach	Device	Qualifier
9 Hip Joint, Right B Hip Joint, Left	0 Open	<u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	5 Metal on Polyethylene 6 Metal on Metal 7 Ceramic on Ceramic 8 Ceramic on Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
A Hip Joint, Acetabular Surface, Right E Hip Joint, Acetabular Surface, Left	0 Open	<u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	F Metal G Ceramic H Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
R Hip Joint, Femoral Surface, Right S Hip Joint, Femoral Surface, Left	0 Open	<u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	F Metal G Ceramic Z No Qualifier

Option 3: Add two new device values, one specifying cemented implants and one specifying uncemented implants (64 codes). This allows an “unspecified” option if the documentation is insufficient to determine whether the synthetic joint substitute was cemented or uncemented.

Section: 0 Medical and Surgical

Body System: S Lower Joints

Operation: R Replacement: Putting in or on biological or synthetic material that physically takes the place and /or function of all or a portion of a body part

Body Part	Approach	Device	Qualifier
C Knee Joint, Right D Knee Joint, Left F Ankle Joint, Right G Ankle Joint, Left T Knee Joint, Femoral Surface, Right U Knee Joint, Femoral Surface, Left V Knee Joint, Tibial Surface, Right W Knee Joint, Tibial Surface, Left	0 Open	<u>G Synthetic Substitute, Uncemented</u> <u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	Z No Qualifier

Body Part	Approach	Device	Qualifier
9 Hip Joint, Right B Hip Joint, Left	0 Open	<u>G Synthetic Substitute, Uncemented</u> <u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	5 Metal on Polyethylene 6 Metal on Metal 7 Ceramic on Ceramic 8 Ceramic on Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
A Hip Joint, Acetabular Surface, Right E Hip Joint, Acetabular Surface, Left	0 Open	<u>G Synthetic Substitute, Uncemented</u> <u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	F Metal G Ceramic H Polyethylene Z No Qualifier

Body Part	Approach	Device	Qualifier
R Hip Joint, Femoral Surface, Right S Hip Joint, Femoral Surface, Left	0 Open	<u>G Synthetic Substitute, Uncemented</u> <u>H Synthetic Substitute, Cemented</u> J Synthetic Substitute	F Metal G Ceramic Z No Qualifier

CMS recommendation: Option 3.

Interspinous Process Internal Fixation Procedures: Request for Adding New Values in PCS Tables

The proposal is to add PCS values to the PCS Insertion tables that will distinguish between dynamic stabilization and static distraction interspinous process internal fixation devices of the spine. PCS has standardized values in the Insertion tables for Upper and Lower Joints body systems for insertion of interspinous process devices, without differentiating between dynamic stabilization and static distraction devices. To extend the level of detail available for the root operation Insertion to differentiate the procedures described above requires new qualifier values in the PCS tables as specified in the proposal.

Option 1: Current PCS coding, shown in example OSH Lower Joints body system

Section: 0 Medical and Surgical
 Body System: S Lower Joints
 Operation: H Insertion: Putting in a nonbiological appliance that monitors, assists, performs, or prevents a physiological function but does not physically take the place of a body part

Body Part	Approach	Device	Qualifier
0 Lumbar Vertebral Joint	0 Open	4 Internal Fixation De	2 Interspinous Process
3 Lumbosacral Joint	3 Percutaneous		3 Pedicle-Based Dynar
	4 Percutaneous Endosco		Stabilization
			Z No Qualifier

Option 2: Create new PCS values in tables 0RH and 0SH for vertebral joint body parts to specify Interspinous Process, Dynamic Stabilization and Interspinous Process, Static Distraction in the 7th character. (12 codes)

Section: 0 Medical and Surgical
 Body System: S Lower Joints
 Operation: H Insertion: Putting in a nonbiological appliance that monitors, assists, performs, or prevents a physiological function but does not physically take the place of a body part



Body Part	Approach	Device	Qualifier
0 Lumbar Vertebral Joint 3 Lumbosacral Joint	0 Open 3 Percutaneous 4 Percutaneous Endoscopy	4 Internal Fixation Device	2 Interspinous Process 3 Pedicle-Based Dynamic Stabilization <u>9 Interspinous Process, Dynamic Stabilization</u> <u>B Interspinous Process, Static Distraction</u> Z No Qualifier

Option 3: Revise existing PCS 7th character qualifier value Interspinous Process in tables 0RH and 0SH for vertebral joint body parts to specify Interspinous Process, Dynamic Stabilization. Create new 7th character qualifier value PCS value in tables 0RH and 0SH for vertebral joint body parts to specify Interspinous Process, Static Distraction. (6 codes)

Section: 0 Medical and Surgical

Body System: S Lower Joints

Operation: H Insertion: Putting in a nonbiological appliance that monitors, assists, performs, or prevents a physiological function but does not physically take the place of a body part

Body Part	Approach	Device	Qualifier
0 Lumbar Vertebral Joint 3 Lumbosacral Joint	0 Open 3 Percutaneous 4 Percutaneous Endoscopy	4 Internal Fixation Device	2 Interspinous Process, <u>Dynamic Stabilization</u> 3 Pedicle-Based Dynamic Stabilization <u>9 Interspinous Process, Static Distraction</u> Z No Qualifier

CMS Recommendation: CMS does not have a formal recommendation at this time. We would like to hear comments from the audience regarding documentation for this level of detail.

Proposed Change to Spinal Fusion Procedures: Request for Deleting Device Value in PCS Tables

The proposal is to delete PCS device values from the PCS Fusion tables for the intervertebral joint body parts that specify internal fixation device as the means of accomplishing the fusion of the spine. The rationale is as follows: PCS guidelines indicate the device value for a fusion procedure should be to code to the primary technique for accomplishing the fusion. Clinical opinion is that internal fixation is never the primary technique, and therefore offering this choice may result in fusion procedures being coded incorrectly. (99 codes)

Section: 0 Medical and Surgical
 Body System: R Upper Joints
 Operation: G Fusion: Joining together portions of an articular body part rendering the articular body part immobile

Body Part	Approach	Device	Qualifier
0 Occipital-cervical Joint 1 Cervical Vertebral Joint 2 Cervical Vertebral Joints, 2 or more 4 Cervicothoracic Vertebral Joint 6 Thoracic Vertebral Joint 7 Thoracic Vertebral Joints, 2 to 7 8 Thoracic Vertebral Joints, 8 or more A Thoracolumbar Vertebral Joint	0 Open 3 Percutaneous 4 Percutaneous Endoscopic	3 Interbody Fusion Device 4 Internal Fixation Device 7 Autologous Tissue Substitute J Synthetic Substitute K Nonautologous Tissue Substitute Z No Device	0 Anterior Approach, Anterior Column 1 Posterior Approach, Posterior Column J Posterior Approach, Anterior Column

Section: 0 Medical and Surgical
 Body System: S Lower Joints
 Operation: G Fusion: Joining together portions of an articular body part rendering the articular body part immobile

Body Part	Approach	Device	Qualifier
0 Lumbar Vertebral Joint 1 Lumbar Vertebral Joints, 2 or more 3 Lumbosacral Joint	0 Open 3 Percutaneous 4 Percutaneous Endoscopic	3 Interbody Fusion Device 4 Internal Fixation Device 7 Autologous Tissue Substitute J Synthetic Substitute K Nonautologous Tissue Substitute Z No Device	0 Anterior Approach, Anterior Column 1 Posterior Approach, Posterior Column J Posterior Approach, Anterior Column

CMS Recommendation: CMS supports this proposal. Currently, in ICD-9-CM we do not code internal fixation (instrumentation) separately. We believe this same principle should be



followed in ICD-10-PCS, thereby making it appropriate to delete the Internal Fixation Device option from the PCS code tables.

Implantable Meshes—Request for New Detail in ICD-10-PCS Tables

The proposal is to create PCS values that differentiate between nonautologous tissue of human origin and nonautologous tissue of non-human origin for the root operation Supplement in the Anatomical Regions, Urinary and Female Reproductive body systems.

Reconstruction of soft tissue defects can be performed by suture but, because it reduces tension and decreases the rate of defect recurrence, mesh is used to reconstruct larger defects, for reinforcement and to strengthen tissue.

Mesh is either synthetic or biologic. Synthetic patches and plugs are manufactured from materials such as polypropylene and PTFE. They are easy to use but have a higher risk of complications like infection and adhesions. Biologic mesh is composed of nonautologous tissue. It tends to be used in inpatient cases where patients have aggravating factors such as obesity, heavy smoking, contaminated surgical site, and loss of fascia. Nonautologous mesh is bio-engineered to remove cells which can cause rejection, rendering a collagen-rich matrix. This promotes tissue regeneration and revascularization, reducing the risk of complications.

Nonautologous mesh is either of human origin using tissue derived from cadavers, eg. Alloderm, or of non-human origin (zooplastic) using bovine or porcine tissue, eg. Permacol. There are currently no significant data sources which enable clinical analysis of the differences in patient outcomes, effectiveness and adverse events between nonautologous mesh of human and non-human origin.

PCS has standardized values for distinguishing between nonautologous tissue of human origin and nonautologous tissue of non-human origin, as shown in this excerpt from table 02U, root operation Supplement for the Heart and Great Vessels body system.

Section: 0 Medical and Surgical
 Body System: 2 Heart and Great Vessels
 Operation: U Supplement: Putting in or on biological or synthetic material that physically reinforces and/or augments the function of a portion of a body part

Body Part	Approach	Device	Qualifier
5 Atrial Septum	0 Open	7 Autologous Tissue	Z No Qualifier
6 Atrium, Right	3 Percutaneous	Substitute	
7 Atrium, Left	4 Percutaneous	8 Zooplastic Tissue	
	Endoscopic	J Synthetic Substitute	
		K Nonautologous Tissue Substitute	

To extend the level of detail available for the root operation Supplement to additional body systems requires only adding the existing device value Zooplastic Tissue to the body systems and

root operations specified in the proposal. Example shown is an excerpt from table 0YU, the Lower Extremities Anatomical Regions.

Option 1: Current PCS coding

Section: 0 Medical and Surgical
 Body System: Y Anatomical Regions, Lower Extremities
 Operation: U Supplement: Putting in or on biological or synthetic material that physically reinforces and/or augments the function of a portion of a body part

Body Part	Approach	Device	Qualifier
5 Inguinal Region, Right 6 Inguinal Region, Left 7 Femoral Region, Right 8 Femoral Region, Left A Inguinal Region, Bilateral E Femoral Region, Bilateral	0 Open 4 Percutaneous Endoscopic	7 Autologous Tissue Substitute J Synthetic Substitute K Nonautologous Tissue Substitute	Z No Qualifier

Option 2: Extend existing device value Zooplasic Tissue to root operation Supplement tables in body systems T Urinary, U Female Reproductive, W General Anatomical Regions, and Y Lower Extremities Anatomical Regions. (144 new codes, for all body parts in tables 0TU, 0UU, 0WU and 0YU)

Section: 0 Medical and Surgical
 Body System: Y Anatomical Regions, Lower Extremities
 Operation: U Supplement: Putting in or on biological or synthetic material that physically reinforces and/or augments the function of a portion of a body part

Body Part	Approach	Device	Qualifier
5 Inguinal Region, Right 6 Inguinal Region, Left 7 Femoral Region, Right 8 Femoral Region, Left	0 Open 4 Percutaneous Endoscopic	7 Autologous Tissue Substitute <u>8 Zooplasic Tissue</u> J Synthetic Substitute K Nonautologous Tissue Substitute	Z No Qualifier



Body Part	Approach	Device	Qualifier
A Inguinal Region, Bilateral E Inguinal Region, Bilateral			

CMS recommendation: Option 2.

Intraoperative Nerve Measurement and Monitoring: Request for Extension of Existing Values in PCS Tables

The proposal is to add PCS values to the PCS Measurement and Monitoring section that will expand the body parts and approaches available for coding monitoring procedures of the nervous system. The proposed change would add the body part value Peripheral Nervous and the approach value Percutaneous to the PCS tables 4A0 and 4A1, as specified in the proposal.

Option 1: Current PCS coding, as shown in the root operation Monitoring table 4A1.

Section: 4 Measurement and Monitoring
 Body System: A Physiological Systems
 Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
0 Central Nervous	0 Open X External	4 Electrical Activity	G Intraoperative Z No Qualifier

Option 2: Extend existing PCS values in table 4A0 and 4A1 to include the body part value Peripheral Nervous and the approach value Percutaneous. (12 codes)

Section: 4 Measurement and Monitoring
 Body System: A Physiological Systems
 Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
0 Central Nervous <u>1 Peripheral Nervous</u>	0 Open <u>3 Percutaneous</u> X External	4 Electrical Activity	G Intraoperative Z No Qualifier

CMS Recommendation: Option 2 as displayed above.



Regional Brain Oxygen Saturation Monitoring Using Near-Infrared Spectroscopy: Request for Extension of Existing Values in PCS Tables

The proposal is to add PCS values to the PCS Monitoring tables that will distinguish invasive intracranial oxygen saturation monitoring from non-invasive external monitoring of regional brain oxygen saturation using near-infrared spectroscopy. PCS has standardized values in the Measurement and Monitoring section for recording monitoring of oxygen saturation.

To extend the level of detail available for the root operation Monitoring to differentiate the procedures specified above requires only adding the existing approach value External and the existing qualifier value No Qualifier to the PCS table 4A1, as specified in the proposal.

Option 1: Current PCS coding

Section: 4 Measurement and Monitoring

Body System: A Physiological Systems

Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
0 Central Nervous	3 Percutaneous 7 Via Natural or Artificial Opening	B Pressure K Temperature R Saturation	D Intracranial

Option 2: Extend existing PCS values in table 4A1 to specify External approach in the 5th character and No Qualifier in the 7th character, for external oxygen saturation monitoring as well as pressure and temperature monitoring, for both the regional brain and intracranial sites. (6 codes)

Section: 4 Measurement and Monitoring

Body System: A Physiological Systems

Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
0 Central Nervous	3 Percutaneous 7 Via Natural or Artificial Opening <u>X External</u>	B Pressure K Temperature R Saturation	D Intracranial <u>Z No Qualifier</u>

Option 3: Extend existing PCS values in table 4A1 to specify External approach in the 5th character and No Qualifier in the 7th character, only apply the 6th character function Saturation, to

specify both regional brain and intracranial oxygen saturation monitoring using an external approach. (2 codes)

Section: 4 Measurement and Monitoring

Body System: A Physiological Systems

Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
0 Central Nervous	3 Percutaneous 7 Via Natural or Artificial Opening <u>X External</u>	R Saturation	D Intracranial <u>Z No Qualifier</u>

Regional Somatic Saturation Monitoring Using Near-Infrared Spectroscopy: Request for New Values in PCS Tables

The proposal is to add PCS values to the PCS Monitoring tables to create unique codes for monitoring oxygen saturation in the soft tissue of various body regions using near-infrared spectroscopy; namely, gastrointestinal, skin and breast, subcutaneous tissue, muscle, and kidney. PCS has standardized values in the Measurement and Monitoring section for recording monitoring of oxygen saturation, and several body system values corresponding to some of the body regions specified in the proposal.

To extend the level of detail available for the root operation Monitoring to create unique codes for the procedures specified in the proposal requires adding the existing approach value External in the 5th character and the existing function value Saturation in the 6th character, as well as creating new 4th character physiological system values as needed.

Option 1: Do not add new PCS values to create unique codes.

Option 2: Extend existing PCS values in table 4A1 to specify External approach in the 5th character and Saturation in the 6th character, for non-invasive NIRS oxygen saturation monitoring of existing body regions Gastrointestinal, Urinary, and Musculoskeletal. (5 codes)

Section: 4 Measurement and Monitoring
 Body System: A Physiological Systems
 Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
B Gastrointestinal D Urinary F Musculoskeletal	<u>X External</u>	<u>R Saturation</u>	Z No Qualifier

Create new PCS physiological system values for Skin and Breast, and for Subcutaneous Tissue, and add existing PCS values in table 4A1 to specify External approach in the 5th character and Saturation in the 6th character, for non-invasive NIRS oxygen saturation monitoring, for the skin/breast and subcutaneous tissue body regions.

Section: 4 Measurement and Monitoring
 Body System: A Physiological Systems
 Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
<u>K Skin and Breast</u> <u>L Subcutaneous Tis</u>	<u>X External</u>	<u>R Saturation</u>	<u>Z No Qualifier</u>

Option 3: Create a single new physiological system value Anatomical Regions, and extend existing PCS values in table 4A1 to specify External approach in the 5th character and Saturation in the 6th character, for non-invasive NIRS oxygen saturation monitoring of all anatomical regions other than the brain and circulatory system. (1 code)

Section: 4 Measurement and Monitoring

Body System: A Physiological Systems

Operation: 1 Monitoring: Determining the level of a physiological or physical function repetitively over a period of time

Body System	Approach	Function/Device	Qualifier
<u>W Anatomical Regi</u> <u>General</u>	X External	R Saturation	Z No Qualifier